



POLICE DEPARTMENT

**LAKE HAVASU CITY**

2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.


Michael Maloy Jr. #281  
Signature

STATE OF ARIZONA       )  
  ) ss  
COUNTY OF MOHAVE    )

Subscribed and sworn to before me this 2<sup>nd</sup> day of September  
2025 by Michael Maloy

Jennifer Lentz  
NOTARY PUBLIC

09/02/2025  
DATE



Commission expires: 03/15/2029

Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

E-mail: police@lhcaz.gov

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasu Silent Witness ..... 854-TIPS

281

**EXHIBIT I-2**  
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY**  
**INTOXILYZER MODEL 9000**

**PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE**

QA SPECIALIST M. MALOY #281 AGENCY LHCPD  
DATE 09-02-2025 TIME 01:34  
INTOXILYZER SERIAL # 90-003695

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

**DIAGNOSTIC TESTS**

- ☒ 1. Clock time check  
☒ 2. Date check

**OPERATIONAL TESTS**

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- ☒ 5. Standard Calibration Check:  
0. 098 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

**COMMENTS:**

INSTRUMENT TAKEN OUT OF SERVICE FOR P.S.I.: 300  
ANNUAL P.M.

SIGNATURE

Michael Maloy #281

## ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
 Serial Number: 90-003695  
 Software Version: 9439.01.00  
 Analytical Report Number: 36950902250360

QAS: MALOY, MICHAEL

QAS Permit #: 30178  
 Agency: LAKE HAVASU PD  
 Last 31-Day Check: 08/25/2025  
 Last Annual Maintenance: 10/28/2024

SUBJECT INFORMATION

Name: 31-Day Check  
 Test Date: 09/02/2025  
 Driver's License #:  
 Gender:  
 Date of Birth:  
 Age:

Weight:  
 State of Issue:  
 Driver's License Expiration:  
 Deprivation Start Time: 01:00  
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL  
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100  
 Standard Lot #: 302-403183349  
 Expiration Date: 01/24/2027  
 Bottle #: 060  
 Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	01:34:21
Diagnostic Test	PASS	01:34:56
Air Blank	0.000	01:35:33
Calibration Chk	0.098	01:35:55
Air Blank	0.000	01:36:35
Subject Test 1	DEF*	01:39:55
Air Blank	0.000	01:40:46
Air Blank	0.000	01:41:26
Subject Test 2	0.000	01:41:57
Air Blank	0.000	01:42:38
Wait		01:46:42
Air Blank	0.000	01:47:19
Subject Test 3	INV**	01:47:55
Air Blank	RFI***	01:48:10

RESULTSEXCEPTION MESSAGES

\* Deficient Sample  
 \*\* Invalid Sample - Begin new deprivation period  
 \*\*\* RFI Detect

OPERATOR COMMENTS

\*Deficient Sample  
 \*\*Invalid Sample - Begin new deprivation period  
 \*\*\*RFI Detect

All air blank results must be 0.000.

Consecutive subject test results must not differ by more than 0.020 g/210L.

Standard check results must be  $\pm 10\%$ .

Not a Successfully Completed Test Sequence