



POLICE DEPARTMENT

**LAKE HAVASU CITY**

2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr. #281  
Signature


STATE OF ARIZONA       )  
                                          ) ss  
COUNTY OF MOHAVE    )

Subscribed and sworn to before me this 25 day of AUGUST

2025 by MICHAEL MALOY

[Signature]  
NOTARY PUBLIC

8-25-2025  
DATE



Commission expires: 3-16-2029

Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

E-mail: police@lhcaz.gov

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasus Silent Witness ..... 854-TIPS

## EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

## PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD  
DATE 08-25-2025 TIME 19:39  
INTOXILYZER SERIAL # 90-003679

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.109 AC

## DIAGNOSTIC TESTS

- ☒ 1. Clock time check  
☒ 2. Date check

## OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- ☒ 5. Standard Calibration Check:  
0. 098 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

## COMMENTS:

INSTRUMENT SENT TO CENTRAL P.S.I.: 850  
FOR ANNUAL P.M.

SIGNATURE

Michael Maloy Jr #281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003679  
Software Version: 9439.01.00  
Analytical Report Number: 36790825250301

QAS: MALOY, MICHAEL  
  
QAS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 07/08/2025  
Last Annual Maintenance: 08/14/2024

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 08/25/2025  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 19:15  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
Agency: LAKE HAVASU PD

Permit #: 30178

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-403183349  
Expiration Date: 01/24/2027  
Bottle #: 048  
Last Changed By: MALOY, MICHAEL

Permit #: 30178

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 19:39:49 |
| Diagnostic Test | PASS   | 19:40:25 |
| Air Blank       | 0.000  | 19:41:02 |
| Calibration Chk | 0.098  | 19:41:23 |
| Air Blank       | 0.000  | 19:42:03 |
| Subject Test 1  | DEF*   | 19:45:25 |
| Air Blank       | 0.000  | 19:46:10 |
| Air Blank       | 0.000  | 19:46:51 |
| Subject Test 2  | 0.000  | 19:47:16 |
| Air Blank       | 0.000  | 19:49:00 |
| Wait            |        | 19:53:05 |
| Air Blank       | 0.000  | 19:53:42 |
| Subject Test 3  | INV**  | 19:54:25 |
| Air Blank       | RFI*** | 19:54:40 |

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .