



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.


Michael Maloy Jr. #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 16th day of DECEMBER
2025 by MICHAEL MALOY

Jennifer Lentz
NOTARY PUBLIC

12/16/2025
DATE



Commission expires: 03/15/2029

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasus Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LKCPD
DATE 12-16-2025 TIME 01:16
INTOXILYZER SERIAL # 90-003635

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0. 096 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

P.S.I: 800

STANDARD CHANGE

SIGNATURE



Intoxilyzer - Alcohol Analyzer
Model 9000

281

SN: 90-003695

Test Date: 12/16/2025

Test Time: 01:12

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402972207

Standard Value: 0.100

Expiration Date: 02/20/2026

Bottle #: 024

Last Changed By: MALOY, MICHAEL

Permit #: 30178

Operator: MALOY, MICHAEL

Permit #: 30178

Agency: LAKE HAVASU PD

Test	g/210L	Time
Air Blank	0.000	01:12:48
Calibration Chk	0.096	01:13:09
Air Blank	0.000	01:13:49

No RFI Detected

Successfully Completed Test Sequence

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
 Serial Number: 90-003695
 Software Version: 9439.01.00
 Analytical Report Number: 36951216250385

QAS: MALOY, MICHAEL
 QAS Permit #: 30178
 Agency: LAKE HAVASU PD
 Last 31-Day Check: 11/24/2025
 Last Annual Maintenance: 10/08/2025

SUBJECT INFORMATION

Name: 31-Day Check
 Test Date: 12/16/2025
 Driver's License #:
 Gender:
 Date of Birth:
 Age:

Weight:
 State of Issue:
 Driver's License Expiration:
 Deprivation Start Time: 01:00
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
 Standard Lot #: 302-402972207
 Expiration Date: 02/20/2026
 Bottle #: 024
 Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	01:16:27
Diagnostic Test	PASS	01:17:02
Air Blank	0.000	01:17:40
Calibration Chk	0.096	01:18:02
Air Blank	0.000	01:18:41
Subject Test 1	DEF*	01:21:54
Air Blank	0.000	01:22:35
Air Blank	0.000	01:23:15
Subject Test 2	0.000	01:23:53
Air Blank	0.000	01:24:49
Wait		01:28:53
Air Blank	0.000	01:29:30
Subject Test 3	INV**	01:30:25
Air Blank	RFI***	01:30:38

*Deficient Sample
 **Invalid Sample - Begin new deprivation period
 ***RFI Detect

RESULTSEXCEPTION MESSAGES

* Deficient Sample
 ** Invalid Sample - Begin new deprivation period
 *** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
 Consecutive subject test results must not differ by more than 0.020 g/210L.
 Standard check results must be $\pm 10\%$.