

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST L. ALLEN #251 AGENCY LHC PD

DATE 03-11-2020 TIME 0134

INTOXILYZER SERIAL # 80-003356 LOCATION LHC PD SIGN #8 CHANNEL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #251

#251

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
03/11/2020 01:34:52

Standard Lot#: 13518100A1
Last Changed By: J. SONDRON #187

QAS: L. ALLEN #251
LAKE HAVASU CITY PD

Operator: L. ALLEN #251
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:35:46
0.100 Cal Check	0.097	01:36:06
Air Blank	0.000	01:36:36

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST L. ALLEN #251 AGENCY LHC PD
DATE 03-11-2020 TIME 0140
INTOXILYZER SERIAL # 80-003356 LOCATION LHC PD SIGN #8 CHANNEL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE  #251

LA
#251

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
03/11/2020 01:40:21

Standard Lot#: 13518100A1
Last Changed By: J. SONDRON #187

QAS: L. ALLEN #251
LAKE HAVASU CITY PD

Operator: L. ALLEN #251
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1995
Sex: M Weight: 170

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:41:22
Diagnostic Test	Pass	01:41:57
Air Blank	0.000	01:42:25
0.100 Cal Check	0.097	01:42:45
Air Blank	0.000	01:43:15
Subject Test	0.000	01:43:40
Air Blank	ABT*	01:43:49
Air Blank	IPS**	01:43:57

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

#251

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
03/11/2020

01:44:28

Standard Lot#: 13518100A1
Last Changed By: J. SONDROL #187

QAS: L. ALLEN #251
LAKE HAVASU CITY PD

Operator: L. ALLEN #251
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1995
Sex: M Weight: 170

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:45:27
Diagnostic Test	Pass	01:46:02
Air Blank	0.000	01:46:30
0.100 Cal Check	0.098	01:46:50
Air Blank	0.000	01:47:20
Subject Test	0.000*	01:50:28
Air Blank	0.000	01:50:55
Air Blank	0.000	01:51:22
Subject Test	XXX**	01:51:44
Air Blank	RFI***	01:51:55

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

✓
#251

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST L. ALLEN #251 AGENCY LHC PD


DATE 03-11-2020 TIME 0153

INTOXILYZER SERIAL # 80-003356 LOCATION LHC PD SIGN #8 CHANNEL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #251

LA
#251

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
03/11/2020 01:53:51

Standard Lot#: 13518100A1
Last Changed By: J. SONDROL #187

QAS: L. ALLEN #251
LAKE HAVASU CITY PD

Operator: L. ALLEN #251
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:54:23
0.100 Cal Check	0.097	01:54:44
Air Blank	0.000	01:55:13

Preventative Maintenance
Performed -- Timer Reset