POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 13 day of MARCH.

2024 by MICHAEL MALOY

NOTARY PUBLIC

03/13/2025
DATE

Commission expires: 3/15/2025

JENNIFER LEA LENTZ
Notary Public
Commission Number 601036
Expires: March 15, 2025
Mohave County

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

281

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 03-13-2024 TIME 09:44 HRS
INTOXILYZER SERIAL # 90-003695

☑ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS
☑ 1. Clock time check
☑ 2. Date check

OPERATIONAL TESTS
☑ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☑ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
☑ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☑ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☑ 5. Standard Calibration Check:
0.097 AC
☑ 6. Air Blanks Completed
☑ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS:
PLACED INTO SERVICE SIGN'D P.S.I. 950
STANDARD CHANGE

SIGNATURE Michael Maloy Jr #281

DPS Form Exh I-2 (Rev 19-03)

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003695
Software Version: 9439.01.00
Analytical Report Number: 36950313240211

QAS: MALOY, MICHAEL

QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 10/14/2023
Last Annual Maintenance: 11/02/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 03/13/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 09:25
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402801597
Expiration Date: 03/15/2025
Bottle #: 001
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	09:44:17
Diagnostic Test	PASS	09:44:53
Air Blank	0.000	09:45:30
Calibration Chk	0.097	09:45:52
Air Blank	0.000	09:46:31
Subject Test 1	DEF*	09:49:49
Air Blank	0.000	09:50:40
Air Blank	0.000	09:51:20
Subject Test 2	0.000	09:51:52
Air Blank	0.000	09:52:37
Wait		09:56:41
Air Blank	0.000	09:57:18
Subject Test 3	INV**	09:57:50
Air Blank	RFI***	09:58:03

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.

281

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003695
Test Date: 03/13/2024 Test Time: 09:40
Location:

Software Version: 9439.01.00

Standard Lot #: 302-402801597
Standard Value: 0.100
Expiration Date: 03/15/2025
Bottle #: 001
Last Changed By: MALOY, MICHAEL
Permit #: 30178

Operator: MALOY, MICHAEL
Permit #: 30178
Agency: LAKE HAVASU PD

Test	g/210L	Time
Air Blank	0.000	09:40:48
Calibration Chk	0.098	09:41:10
Air Blank	0.000	09:41:49

No RFI Detected

Successfully Completed Test Sequence