

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDRELL 187 AGENCY LNC PD

DATE 4/5/16 TIME 0429

INTOXILYZER SERIAL # 80-001873 LOCATION LNC MAIN JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record/ Attach the record to the completed checklist.

SIGNATURE J Sandrell 187

DPS Form Exh G-2 (Rev 05-01)

* SAW THIS INTOX HAD BEEN SWITCHED OUT BY AUVIN OF QAS TRELSCHER.
* DIDN'T REALIZE HE DID NOT COMPLETE A 31+90 DAY AFTER PUTTING INTO SERVICE.

INTOXILYZER 8000

Location: LHC MAIN JAIL

Serial Number: 80-001873

Core Version: 8105.48

04/05/2016

04:29:44

Standard Lot#: 22315100A1

Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187

LAKE HAVASU CITY PD

Operator: J. SONDROL #187

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	04:30:19
0.100 Cal Check	0.100	04:30:40
Air Blank	0.000	04:31:10

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SANDRILL 187 AGENCY LNC PD

DATE 4/5/16 TIME 0432

INTOXILYZER SERIAL # 80-001873 LOCATION LNC MAIN JAIL

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE J Sandrill 187

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
04/05/2016

04:32:58

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:33:45
Diagnostic Test	Pass	04:34:20
Air Blank	0.000	04:34:48
0.100 Cal Check	0.100	04:35:08
Air Blank	0.000	04:35:38
Subject Test	0.000	04:35:56
Air Blank	ABT*	04:36:03
Air Blank	IPS**	04:36:10

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
04/05/2016

04:37:11

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:38:01
Diagnostic Test	Pass	04:38:36
Air Blank	0.000	04:39:03
0.100 Cal Check	0.099	04:39:24
Air Blank	0.000	04:39:54
Subject Test	0.145*	04:43:01
Air Blank	0.000	04:43:35
Air Blank	0.000	04:44:02
Subject Test	XXX**	04:44:28
Air Blank	RFI***	04:44:36

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDRAC 187 AGENCY LNC PD

DATE 4/5/16 TIME 0945

INTOXILYZER SERIAL # 80-001873 LOCATION LNC M.A. ~ JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sandrac 187

INTOXILYZER 8000

Location: IHC MAIN JAIL

Serial Number: 80-001873

Core Version: 8105.48

04/05/2016

04:45:26

Standard Lot#: 22315100A1

Last Changed By: J. SONDRON #187

QAS: J. SONDRON #187

LAKE HAVASU CITY PD

Operator: J. SONDRON #187

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	04:45:55
0.100 Cal Check	0.099	04:46:16
Air Blank	0.000	04:46:46

Preventative Maintenance
Performed - Timer Reset