

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESNER #219 AGENCY LAKE HAVASU CITY POLICE

DATE 8-3-2016 TIME 1858 HRS

INTOXILYZER SERIAL # 20-001873 LOCATION LHC RD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

219

INTOXILYZER 8900
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
08/03/2016 18:58:43

Standard Lot#: 22315100A1
Last Changed By: J. SONDRON #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	19:00:08
0.100 Cal Check	0.097	19:00:29
Air Blank	0.000	19:00:59

Preventative Maintenance
Performed -- Timer Reset

ET
*219

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER #219 AGENCY LAKE HAVASU CITY POLICE

DATE 8-3-2015 TIME 1902

INTOXILYZER SERIAL # 80-001873 LOCATION LACPD JAIL

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

AT
219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
08/03/2016

19:02:12

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:03:25
Diagnostic Test	Pass	19:04:00
Air Blank	0.000	19:04:28
0.100 Cal Check	0.097	19:04:49
Air Blank	0.000	19:05:18
Subject Test	0.000	19:05:36
Air Blank	ABT*	19:05:43
Air Blank	IPS**	19:05:51

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

17
219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
08/03/2016 19:02:12

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:03:25
Diagnostic Test	Pass	19:04:00
Air Blank	0.000	19:04:28
0.100 Cal Check	0.097	19:04:49
Air Blank	0.000	19:05:18
Subject Test	0.000	19:05:36
Air Blank	ABT*	19:05:43
Air Blank	IPS**	19:05:51

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

C1
219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80 001873
Core Version: 8105.48
08/03/2016

19:02:12

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:03:25
Diagnostic Test	Pass	19:04:00
Air Blank	0.000	19:04:28
0.100 Cal Check	0.097	19:04:49
Air Blank	0.000	19:05:18
Subject Test	0.000	19:05:36
Air Blank	ABT*	19:05:43
Air Blank	IPS**	19:05:51

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
08/03/2016

19:07:36

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:08:45
Diagnostic Test	Pass	19:09:20
Air Blank	0.000	19:09:47
0.100 Cal Check	0.098	19:10:08
Air Blank	0.000	19:10:38
Subject Test	0.042*	19:13:45
Air Blank	0.000	19:14:16
Air Blank	0.000	19:14:44
Subject Test	XXX**	19:15:04
Air Blank	RFI***	19:15:19

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

CI
217

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
08/03/2016

19:07:36

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:08:45
Diagnostic Test	Pass	19:09:20
Air Blank	0.000	19:09:47
0.100 Cal Check	0.098	19:10:08
Air Blank	0.000	19:10:38
Subject Test	0.042*	19:13:45
Air Blank	0.000	19:14:16
Air Blank	0.000	19:14:44
Subject Test	XXX**	19:15:04
Air Blank	RFI***	19:15:19

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

ET
217

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
08/03/2016

19:07:36

Standard Lot#: 22315100A1
Last Changed By: J. SONDRON #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP, TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:08:45
Diagnostic Test	Pass	19:09:20
Air Blank	0.000	19:09:47
0.100 Cal Check	0.098	19:10:08
Air Blank	0.000	19:10:38
Subject Test	0.042*	19:13:45
Air Blank	0.000	19:14:16
Air Blank	0.000	19:14:44
Subject Test	XXX**	19:15:04
Air Blank	RFI***	19:15:19

*Deficient Sample
**Invalid Sample
 Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

#219

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)


ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER #219 AGENCY LAKE HAVASU CITY POLICE

DATE 8-3-2016 TIME 1916

INTOXILYZER SERIAL # 80-001873 LOCATION LHC PD JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
08/03/2016 19:16:39

Standard Lot#: 22315100A1
Last Changed By: J. SONDRON #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/2'00L	Time
Air Blank	0.000	19:17:10
0.100 Cal Check	0.098	19:17:30
Air Blank	0.000	19:18:00

Preventative Maintenance
Performed Timer Reset