



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Josh Sondrol
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 24 day of February.

2016 by Josh Sondrol
Carla Duarte

NOTARY PUBLIC

2-24-16
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
 TDD 855-4114
Havasus Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRAL 187 AGENCY LNC PD
DATE 2/23/16 TIME 2317
INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondral 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
02/23/2016 23:17:50

Standard Lots: 22315100R1
Last Changed By: J. SONDRAL #187

DAS: J. SONDRAL #187
LAKE HAVASU CITY PD

Operator: J. SONDRAL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	23:18:58
0.100 Cal Check	0.100	23:19:18
Air Blank	0.000	23:19:48

Preventative Maintenance
Performed -- Timer Reset

9187

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD
DATE 2/23/16 TIME 2322
INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

J Sondrol 187

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
02/23/2016 23:29:03

Standard Lot#: 22315100R1
Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:28:51
Diagnostic Test	Pass	23:29:29
Air Blank	0.000	23:29:56
0.100 Cal Check	0.100	23:30:17
Air Blank	0.000	23:30:47
Subject Test	0.270*	23:33:54
Air Blank	0.000	23:34:27
Air Blank	0.000	23:34:54
Subject Test	XXX**	23:35:22
Air Blank	RF***	23:35:31

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RF: Detect

INTOXILYZER 8110
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
02/23/2016 23:22:29

Standard Lot#: 22315100R1
Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:23:24
Diagnostic Test	Pass	23:24:02
Air Blank	0.000	23:24:29
0.100 Cal Check	0.100	23:24:50
Air Blank	0.000	23:25:19
Subject Test	0.000	23:25:42
Air Blank	RF*	23:25:58
Air Blank	1P5**	23:25:57

*Sequence Aborted
**Improper Sample

J Sondrol

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD
DATE 2/23/16 TIME 2336
INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
02/23/2016 23:36:39

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

DPS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	23:37:08
0.100 Cal Check	0.098	23:37:29
Air Blank	0.000	23:37:58

Preventative Maintenance
Performed -- Timer Reset