



POLICE DEPARTMENT

LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 26th day of April.

2023 by Michael Maloy Jr

[Signature]
NOTARY PUBLIC

4-26-23
DATE 2-26-27



Commission expires:

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPA
DATE 04-26-2023 TIME 07:17
INTOXILYZER SERIAL # 90-003695

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0. 103 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

STANDARD CHANGE

SIGNATURE

Michael Maloy Jr #281

281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
 Serial Number: 90-003695
 Software Version: 9439.01.00
 Analytical Report Number: 36950426230117

QAS: MALOY, MICHAEL
 QAS Permit #: 30178
 Agency: LAKE HAVASU PD
 Last 31-Day Check: 03/30/2023
 Last Annual Maintenance: 11/28/2022

SUBJECT INFORMATION

Name: 31-Day Check
 Test Date: 04/26/2023
 Driver's License #:
 Gender:
 Date of Birth:
 Age:

Weight:
 State of Issue:
 Driver's License Expiration:
 Deprivation Start Time: 07:00
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
 Standard Lot #: 302-402454776
 Expiration Date: 06/16/2024
 Bottle #: 001
 Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	07:17:20
Diagnostic Test	PASS	07:17:56
Air Blank	0.000	07:18:33
Calibration Chk	0.103	07:18:54
Air Blank	0.000	07:19:34
Subject Test 1	DEF*	07:22:50
Air Blank	0.000	07:23:32
Air Blank	0.000	07:24:13
Subject Test 2	0.000	07:24:39
Air Blank	0.000	07:25:21
Wait		07:29:26
Air Blank	0.000	07:30:03
Subject Test 3	INV**	07:30:27
Air Blank	RFI***	07:30:48

RESULTSEXCEPTION MESSAGES

*Deficient Sample
 **Invalid Sample - Begin new deprivation period
 ***RFI Detect

* Deficient Sample
 ** Invalid Sample - Begin new deprivation period
 *** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
 Consecutive subject test results must not differ by more than 0.020 g/210L.
 Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003695

Test Date: 04/26/2023

Test Time: 07:14

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402454776

Standard Value: 0.100

Expiration Date: 06/16/2024

Bottle #: 001

Last Changed By: MALOY, MICHAEL

Permit #: 30178

Operator: MALOY, MICHAEL

Permit #: 30178

Agency: LAKE HAVASU PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	07:15:20
Calibration Chk	0.102	07:15:42
Air Blank	0.000	07:16:21

No RFI Detected

Successfully Completed Test Sequence