



281

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALBY #281 AGENCY LHC PD  
DATE 04-04-2024 TIME 19:44  
INTOXILYZER SERIAL # 90-003695

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0.100 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

P.S.I: 975

SIGNATURE Michael Malby #281

281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003695  
Software Version: 9439.01.00  
Analytical Report Number: 36950404240212

QAS: MALOY, MICHAEL  
QAS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 03/13/2024  
Last Annual Maintenance: 11/02/2023

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 04/04/2024  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 19:25  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
Agency: LAKE HAVASU PD

Permit #: 30178  
STANDARD INFORMATION  
Standard Value: 0.100  
Standard Lot #: 302-402801597  
Expiration Date: 03/15/2025  
Bottle #: 001  
Last Changed By: MALOY, MICHAEL

Test	g/210L	Time
Air Blank	0.000	19:44:38
Diagnostic Test	PASS	19:45:14
Air Blank	0.000	19:45:51
Calibration Chk	0.100	19:46:13
Air Blank	0.000	19:46:52
Subject Test 1	DEF*	19:50:15
Air Blank	0.000	19:51:02
Air Blank	0.000	19:51:43
Subject Test 2	0.000	19:52:16
Air Blank	0.000	19:52:59
Wait		19:57:04
Air Blank	0.000	19:57:41
Subject Test 3	INV**	19:58:20
Air Blank	RFI***	19:58:34

Permit #: 30178

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*