



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 6 day of APRIL.

2022 by MICHAEL MALOY
Carla Duarte

NOTARY PUBLIC

4-6-22
DATE

Commission expires:



Area Code 928

- | | | | | | |
|-----------------------|----------|-----------------------------|----------|-------------------------------|----------|
| Administration | 855-4884 | Fax #'s | | Emergency | 9-1-1 |
| Business Office | 855-1171 | Administration | 680-5430 | Non-Emergency Dispatch/ | 855-4111 |
| Investigations | 855-5775 | Business Office | 680-5431 | Havasu Silent Witness | 854-TIPS |
| Patrol | 855-0515 | Investigations/Patrol | 680-5432 | | |

E-mail: police@lhcaz.gov

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LNCPO
DATE 04-06-2022 TIME 1619 HRS
INTOXILYZER SERIAL # 90-003680

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.097 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE Michael Maloy Jr #281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
 Serial Number: 90-003680
 Software Version: 9439.01.00
 Analytical Report Number: 36800406220008

QAS: MALOY, MICHAEL
 QAS Permit #: 30178
 Agency: LAKE HAVASU PD
 Last 31-Day Check: 03/03/2022
 Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Name: 31-Day Check
 Test Date: 04/06/2022
 Driver's License #:
 Gender:
 Date of Birth:
 Age:

Weight:
 State of Issue:
 Driver's License Expiration:
 Deprivation Start Time: 16:00
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
 Standard Lot #: 23621100A4
 Expiration Date: 11/05/2023
 Bottle #: 002
 Last Changed By: SLOWINSKI, JOSEPH

Permit #: 10002

Test	g/210L	Time
Air Blank	0.000	16:19:34
Diagnostic Test	PASS	16:20:10
Air Blank	0.000	16:20:47
Calibration Chk	0.097	16:21:09
Air Blank	0.000	16:21:48
Subject Test 1	DEF*	16:25:02
Air Blank	0.000	16:25:49
Air Blank	0.000	16:26:29
Subject Test 2	0.000	16:26:59
Air Blank	0.000	16:27:41
Wait		16:31:45
Air Blank	0.000	16:32:22
Subject Test 3	INV**	16:32:53
Air Blank	RFI***	16:33:06

RESULTSEXCEPTION MESSAGES

*Deficient Sample
 **Invalid Sample - Begin new deprivation period
 ***RFI Detect

* Deficient Sample
 ** Invalid Sample - Begin new deprivation period
 *** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
 Consecutive subject test results must not differ by more than 0.020 g/210L.
 Standard check results must be $\pm 10\%$.