

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Sandoval 187 AGENCY LNC PD
DATE 4/9/19 TIME 1017
INTOXILYZER SERIAL # 80-003356 LOCATION JAIL

(1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

- (2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (4. Air blank completed.
- (5. Calibration check completed. Test results 0.099 AC.
- (6. Air blank completed.
- (7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

J. Sandoval 187

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

* ORIGINAL FORM, FORGOT TO PUT # 5 CAL CHECK RESULTS. J. Sandoval 187