



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281

 Signature

STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)


Subscribed and sworn to before me this 13 day of AUGUST,
2022 by MICHAEL MALOY

[Signature]
 NOTARY PUBLIC

8/13/2022

 DATE

Commission expires: 3-16-2025



Area Code 928

Administration 855-4884
 Business Office 855-1171
 Investigations 855-5775
 Patrol 855-0515

Fax #'s
 Administration 680-5430
 Business Office 680-5431
 Investigations/Patrol 680-5432

Emergency 9-1-1
 Non-Emergency Dispatch/ 855-4111
 Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 08-13-2022 TIME 1932
INTOXILYZER SERIAL # 90-003679

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 107 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE Michael Maloy Jr #281

281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003679
Software Version: 9439.01.00
Analytical Report Number: 36790813220121

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 07/22/2022
Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 08/13/2022
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 19:15
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 23621100A4
Expiration Date: 11/05/2023
Bottle #: 006
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	19:32:03
Diagnostic Test	PASS	19:32:38
Air Blank	0.000	19:33:15
Calibration Chk	0.107	19:33:37
Air Blank	0.000	19:34:16
Subject Test 1	DEF*	19:37:33
Air Blank	0.000	19:38:20
Air Blank	0.000	19:39:00
Subject Test 2	0.000	19:39:34
Air Blank	0.000	19:40:15
Wait		19:44:19
Air Blank	0.000	19:44:56
Subject Test 3	INV**	19:45:25
Air Blank	RFI***	19:45:37

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.