



POLICE DEPARTMENT

LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 15th day of August

2023 by Michael Maloy

Jennifer Le Lentz
NOTARY PUBLIC

8/14/2023
DATE

Commission expires: 3/15/25



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 08-15-2023 TIME 0958
INTOXILYZER SERIAL # 90-003695

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.101 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

PSI 225

SIGNATURE

Michael Maloy Jr #281

281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003695
Software Version: 9439.01.00
Analytical Report Number: 36950815230174

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 07/14/2023
Last Annual Maintenance: 11/28/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 08/15/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 09:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 302-402454776
Expiration Date: 06/16/2024
Bottle #: 001
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	09:58:58
Diagnostic Test	PASS	09:59:34
Air Blank	0.000	10:00:11
Calibration Chk	0.101	10:00:32
Air Blank	0.000	10:01:12
Subject Test 1	DEF*	10:04:30
Air Blank	0.000	10:05:11
Air Blank	0.000	10:05:51
Subject Test 2	0.000	10:06:17
Air Blank	0.000	10:06:58
Wait		10:11:01
Air Blank	0.000	10:11:38
Subject Test 3	INV**	10:12:03
Air Blank	RFI***	10:12:15

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.