



POLICE DEPARTMENT

**LAKE HAVASU CITY**

2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael T. Maloy Jr #261  
Signature

STATE OF ARIZONA           )  
  ) ss  
COUNTY OF MOHAVE       )

Subscribed and sworn to before me this 17<sup>th</sup> day of August.

2023 by Michael T Maloy Jr.

[Signature]  
NOTARY PUBLIC

8-17-23  
DATE

Commission expires: 2-26-27



Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasu Silent Witness ..... 854-TIPS

E-mail: police@lhcaz.gov

281

## EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

## PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO  
DATE 08-16-2023 TIME 1358  
INTOXILYZER SERIAL # 90-003680

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

## DIAGNOSTIC TESTS

- ☒ 1. Clock time check  
☒ 2. Date check

## OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- ☒ 5. Standard Calibration Check:  
0. 097 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

945 PSI

SIGNATURE

Michael Maloy Jr #281

281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003680  
Software Version: 9439.01.00  
Analytical Report Number: 36800816230032

QAS: MALOY, MICHAEL  
QAS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 07/14/2023  
Last Annual Maintenance: 11/29/2022

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 08/16/2023  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 13:30  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
Agency: LAKE HAVASU PD

Permit #: 30178

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-402376471  
Expiration Date: 06/02/2024  
Bottle #: 053  
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	13:58:57
Diagnostic Test	PASS	13:59:32
Air Blank	0.000	14:00:09
Calibration Chk	0.097	14:00:30
Air Blank	0.000	14:01:10
Subject Test 1	DEF*	14:04:26
Air Blank	0.000	14:05:09
Air Blank	0.000	14:05:49
Subject Test 2	0.000	14:06:20
Air Blank	0.000	14:07:08
Wait		14:11:11
Air Blank	0.000	14:11:48
Subject Test 3	INV**	14:12:18
Air Blank	RFI***	14:12:31

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

## RESULTS

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .*