



POLICE DEPARTMENT

**LAKE HAVASU CITY**

2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281  
Signature

STATE OF ARIZONA            )  
  ) ss  
COUNTY OF MOHAVE        )

Subscribed and sworn to before me this 27 day of AUGUST


2024 by MICHAEL MALOY

[Signature]

NOTARY PUBLIC

8-27-2024

DATE



Commission expires: 3-16-2025

Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

E-mail: police@lhcaz.gov

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasus Silent Witness ..... 854-TIPS

281

## EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

## PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALBY #281 AGENCY LUCPD  
DATE 08-27-2024 TIME 00:06  
INTOXILYZER SERIAL # 90-003695

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

## DIAGNOSTIC TESTS

- ☒ 1. Clock time check  
☒ 2. Date check

## OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- ☒ 5. Standard Calibration Check:  
0. 098 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

P.S.I.: 900

SIGNATURE

Michael Malby #281

281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003695  
Software Version: 9439.01.00  
Analytical Report Number: 36950827240218

QAS: MALOY, MICHAEL  
  
QAS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 08/01/2024  
Last Annual Maintenance: 11/02/2023

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 08/27/2024  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 23:45  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
Agency: LAKE HAVASU PD

Permit #: 30178

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-402801597  
Expiration Date: 03/15/2025  
Bottle #: 001  
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	00:06:05
Diagnostic Test	PASS	00:06:40
Air Blank	0.000	00:07:17
Calibration Chk	0.098	00:07:38
Air Blank	0.000	00:08:18
Subject Test 1	DEF*	00:11:35
Air Blank	0.000	00:12:17
Air Blank	0.000	00:12:57
Subject Test 2	0.000	00:13:29
Air Blank	0.000	00:14:10
Wait		00:18:15
Air Blank	0.000	00:18:52
Subject Test 3	INV**	00:19:41
Air Blank	RFI***	00:19:52

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

## RESULTS

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .*