



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
 Signature


STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)

Subscribed and sworn to before me this 11th day of December
2021 by MICHAEL MALOY

Shawna Mcknight
 NOTARY PUBLIC

DEC 11, 2021
 DATE

Commission expires:



Area Code 928

Administration 855-4884	Fax #'s	Emergency 9-1-1
Business Office 855-1171	Administration 680-5430	Non-Emergency Dispatch/ 855-4111
Investigations 855-5775	Business Office 680-5431	Havasu Silent Witness 854-TIPS
Patrol 855-0515	Investigations/Patrol 680-5432	

E-mail: police@lhcaz.gov

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 12-10-2021 TIME 1344
INTOXILYZER SERIAL # 90-663695

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0.100 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE Michael Maloy Jr

ARIZONA FORENSIC BREATH ALCOHOL
ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: INTOXLYZER 9000
Serial Number: 90-003695
Software Version: 9439.01.00
Analytical Report Number: 36951210210003
QAS:
QAS Permit #:
Agency:
Last 31-Day Check:
Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Test Date: 12/10/2021

Name: 31-Day Check,

Driver's License #:

Weight:

Gender:

State of Issue:

Date of Birth:

Driver's License Expiration:

Age:

Deprivation Start Time: 13:23

15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL

Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100

Standard Lot #: 2362100A4

Expiration Date: 11/05/2023

Bottle #: 003

Last Changed By:

SLOWINSKI, JOSEPH

Permit #: 10002

Test	g/210L	Time
Air Blank	0.000	13:44:27
Diagnostic Test	PASS	13:45:02
Air Blank	0.000	13:45:39
Calibration Chk	0.100	13:46:01
Air Blank	0.000	13:46:40
Subject Test 1	DEF*	13:50:01
Air Blank	0.000	13:50:49
Air Blank	0.000	13:51:29
Subject Test 2	0.000	13:51:59
Air Blank	0.000	13:52:48
Wait		13:56:52
Air Blank	0.000	13:57:29
Subject Test 3	INV**	13:58:37
Air Blank	RFI***	13:58:57

*Deficient Sample

**Invalid Sample - Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample

** Invalid Sample - Begin new deprivation period

*** RFI Detect

All air blank results must be 0.000.

Consecutive subject test results must not differ

by more than 0.020 g/210L.

Standard check results must be ±10%.