



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature


STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 7 day of DECEMBER

2022 by MICHAEL MALOY

NOTARY PUBLIC

[Signature]
DATE



Commission expires:

3-16-2025

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO
DATE 12-07-2022 TIME 12:57
INTOXILYZER SERIAL # 90-003679

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0. 100 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

RETURNED FROM A.P.S. FOR YEARLY PREVENTATIVE
MAINTENANCE INSPECTION

SIGNATURE

Michael Maloy Jr #281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003679
Software Version: 9439.01.00
Analytical Report Number: 36791207220123

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 12/07/2022
Last Annual Maintenance: 11/29/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 12/07/2022
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 12:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 1463724
Expiration Date: 03/10/2023
Bottle #: 003
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	12:57:55
Diagnostic Test	PASS	12:58:30
Air Blank	0.000	12:59:07
Calibration Chk	0.100	12:59:29
Air Blank	0.000	13:00:08
Subject Test 1	DEF*	13:03:21
Air Blank	0.000	13:04:06
Air Blank	0.000	13:04:47
Subject Test 2	0.000	13:05:32
Air Blank	0.000	13:06:13
Wait		13:10:17
Air Blank	0.000	13:10:54
Subject Test 3	INV**	13:11:20
Air Blank	RFI***	13:11:33

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.