



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael T. Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 7 day of DECEMBER.

2022 by MICHAEL MALOY

Maloy

NOTARY PUBLIC

12-7-2022

DATE



Commission expires:

3-16-2025

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s

Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

E-mail: police@lhcaz.gov

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 12-07-2022 TIME 17:17
INTOXILYZER SERIAL # 90-003680

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0. 102 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

RETURNED FROM D.P.S. FOR YEARLY PREVENTATIVE

MAINTENANCE INSPECTION.

SIGNATURE

Michael Maloy

ARIZONA FORENSIC BREATH ALCOHOL
ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003680
Software Version: 9439.01.00
Analytical Report Number: 36801207220024
QAS: MALLOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 09/28/2022
Last Annual Maintenance: 11/29/2022

SUBJECT INFORMATION

Test Date: 12/07/2022
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 17:00
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALLOY, MICHAEL
Agency: LAKE HAVASU PD
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402376471
Expiration Date: 06/02/2024
Bottle #: 053
Last Changed By:
MALLOY, MICHAEL
Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	17:17:44
Diagnostic Test	PASS	17:18:20
Air Blank	0.000	17:18:56
Calibration Chk	0.102	17:19:18
Air Blank	0.000	17:19:57
Subject Test 1	DEF*	17:23:21
Air Blank	0.000	17:24:11
Air Blank	0.000	17:24:51
Subject Test 2	0.000	17:25:30
Air Blank	0.000	17:26:26
Wait		17:30:29
Air Blank	0.000	17:31:06
Subject Test 3	INV**	17:31:43
Air Blank	RFI***	17:31:56
*Deficient Sample		
**Invalid Sample - Begin new deprivation period		
***RFI Detect		

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ
by more than 0.020 g/210L.
Standard check results must be ±10%.