



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr # 281
 Signature

STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)

Subscribed and sworn to before me this 29th day of January.

2021 by Michael T. Maloy Jr.

[Signature]
 NOTARY PUBLIC

1-29-21
 DATE

Commission expires: 2-26-23



Area Code 928

Administration 855-4884	Fax #'s	Emergency 9-1-1
Business Office 855-1171	Administration 680-5430	Non-Emergency Dispatch/ 855-4111
Investigations 855-5775	Business Office 680-5431	Havasu Silent Witness 854-TIPS
Patrol 855-0515	Investigations/Patrol 680-5432	

E-mail: police@lhcaz.gov

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO

DATE 01-29-2021 TIME 0754

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPO

- (/) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (/) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (/) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (/) 4. Air blank completed.
- (/) 5. Calibration check completed. Test results 0.100 AC.
- (/) 6. Air blank completed.
- (/) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy #281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-061029
Core Version: 8105.48
01/29/2021 07:54:37

Standard Lot#: 35019100A1
Last Changed By: K. TREBES #7259

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	07:55:08
0.100 Cal Check	0.100	07:55:29
Air Blank	0.000	07:55:59

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 01-29-2021 TIME 0759
INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SENT TO D.P.S. FOR ANNUAL PREVENTATIVE
MAINTENANCE

SIGNATURE Michael Maloy #281

INTOXILYZER 8000
 Location: LHCPD
 Serial Number: 80-001029
 Core Version: 8105.48
 01/29/2021 07:59:46

Standard Lot#: 35019100A1
 Last Changed By: K. TREBES #7259

OAS: M. MALOY JR #281
 LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
 LAKE HAVASU CITY PD

Subject: SQAP,TEST
 DOB: 01/01/2000
 Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:00:36
Diagnostic Test	Pass	08:01:13
Air Blank	0.000	08:01:40
0.100 Cal Check	0.100	08:02:01
Air Blank	0.000	08:02:31
Subject Test	0.000	08:02:45
Air Blank	ABT*	08:02:52
Air Blank	IPS**	08:03:00

*Sequence Aborted
 **Improper Sample

Not a Successfully
 Completed
 Test Sequence

INTOXILYZER 8000
 Location: LHCPD
 Serial Number: 80-001029
 Core Version: 8105.48
 01/29/2021 08:03:24

Standard Lot#: 35019100A1
 Last Changed By: K. TREBES #7259

OAS: M. MALOY JR #281
 LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
 LAKE HAVASU CITY PD

Subject: SQAP,TEST
 DOB: 01/01/2000
 Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:04:14
Diagnostic Test	Pass	08:04:51
Air Blank	0.000	08:05:19
0.100 Cal Check	0.099	08:05:40
Air Blank	0.000	08:06:09
Subject Test	0.000*	08:09:16
Air Blank	0.000	08:09:44
Air Blank	0.000	08:10:11
Subject Test	XXX**	08:10:30
Air Blank	RFI***	08:10:39

*Deficient Sample
 **Invalid Sample
 Begin new deprivation period
 ***RFI Detect

Not a Successfully
 Completed
 Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LUCPD

DATE 01-29-2021 TIME 0811

INTOXILYZER SERIAL # 80-001029 LOCATION LUCPD

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy #281

INTOXILYZER 8000
Location: LMCPO
Serial Number: 80-001029
Core Version: 8105.48
01/29/2021 08:11:12

Standard Lot#: 35019100A1
Last Changed By: K. TREBES #7259

OAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	08:11:42
0.100 Cal Check	0.098	08:12:03
Air Blank	0.000	08:12:32

Preventative Maintenance
Performed -- Timer Reset