



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael T. Maloy Jr #281
 Signature


STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)

Subscribed and sworn to before me this 15th day of January
2021 by Michael T. Maloy Jr

[Signature]
 NOTARY PUBLIC

1-15-21
 DATE

Commission expires:



Area Code 928

- | | | |
|--------------------------------|--------------------------------------|--|
| Administration 855-4884 | Fax #'s | Emergency 9-1-1 |
| Business Office 855-1171 | Administration 680-5430 | Non-Emergency Dispatch/ 855-4111 |
| Investigations 855-5775 | Business Office 680-5431 | Havasu Silent Witness 854-TIPS |
| Patrol 855-0515 | Investigations/Patrol 680-5432 | |

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD

DATE 01-10-2021 TIME 0700

INTOXILYZER SERIAL # 80-001873 LOCATION JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy Jr #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001873
Core Version: 8105.48
01/10/2021 07:00:06

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	07:00:37
0.100 Cal Check	0.098	07:00:58
Air Blank	0.000	07:01:28

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LVLCPD

DATE 01-10-2021 TIME 0702

INTOXILYZER SERIAL # 80-001873 LOCATION JAIL

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

() 1. Clock time check.

() 2. Date check.

OPERATIONAL TESTS

() 1. Alcohol-free subject test result 0. 000 AC.

() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed

() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.

() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Michael Maloy Jr #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 804001873
Core Version: 8105.48
01/10/2021

07:02:46

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:03:35
Diagnostic Test	Pass	07:04:10
Air Blank	0.000	07:04:38
0.100 Cal Check	0.098	07:04:58
Air Blank	0.000	07:05:28
Subject Test	0.000	07:05:58
Air Blank	ABT*	07:06:05
Air Blank	IPS**	07:06:13

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001873
Core Version: 8105.48
01/10/2021

07:06:43

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:07:35
Diagnostic Test	Pass	07:08:10
Air Blank	0.000	07:08:37
0.100 Cal Check	0.098	07:08:58
Air Blank	0.000	07:09:27
Subject Test	0.000*	07:12:35
Air Blank	0.000	07:13:02
Air Blank	0.000	07:13:30
Subject Test	XXX**	07:13:50
Air Blank	RFI***	07:14:17

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LMCPS
DATE 01-10-2021 TIME 0714
INTOXILYZER SERIAL # 86-001873 LOCATION JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy Jr #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001873
Core Version: 8105.48
01/10/2021 07:14:54

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	07:15:24
0.100 Cal Check	0.097	07:15:45
Air Blank	0.000	07:16:15

Preventative Maintenance
Performed -- Timer Reset