

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS M. PLYNNEY III AGENCY LHCRD

DATE 01-21-19 TIME 1011

INTOXILYZER SERIAL # 80-003355 LOCATION LHCRD JAIL

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC. into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed
- 5. Calibration check completed. Test results 0.100 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 263

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: JAIL

Serial Number: 80-003355

Core Version: 8105.43

01/21/2019

10:11:07

Standard Lot#: 13917100A

Last Changed By: J. SONDRON #187

QAS: L. PLUNKETT III #253

LAKE HAVASU CITY PD

Operator: L. PLUNKETT III #253

LAKE HAVASU CITY PD

Test	g/2101	Time
Air Blank	0.000	10:11:38
0.100 Cal Check	0.100	10:11:59
Air Blank	0.000	10:12:28

Preventative Maintenance
Performed -- Timer Reset

Exhibit G-4. Standard Quality Assurance Procedure Intoxilyzer, Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOUIS H. BLUNKETT III AGENCY LHCPD

DATE 01-21-19 TIME 1014

INTOXILYZER SERIAL # 80-003365 LOCATION LHCPD JAIL

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (X) 1. Clock time check.
- (X) 2. Date check.

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0. 000 AC.
- (X) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (X) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (X) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS CLOCK WAS APPROXIMATELY 4 MINUTES SLOW. I CHANGED IT
BEFORE THE FIRST G-2.

SIGNATURE  255

DPS Form Exh G-4 (Rev 05-01)

Historical Note

New Exhibit G-4 made by final rulcmaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
01/21/2019

10:14:14

Standard Lot#: 13917100A1
Last Changed By: J. SONDRONL #187

QAS: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Operator: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:15:16
Diagnostic Test	Pass	10:15:51
Air Blank	0.000	10:16:18
0.100 Cal Check	0.100	10:16:39
Air Blank	0.000	10:17:09
Subject Test	0.000	10:17:41
Air Blank	ABT*	10:17:47
Air Blank	IPS**	10:17:55

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
01/21/2019

10:18:19

Standard Lot#: 13917100A1
Last Changed By: J. SONDROL #187

QAS: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Operator: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:19:04
Diagnostic Test	Pass	10:19:39
Air Blank	0.000	10:20:06
0.100 Cal Check	0.100	10:20:27
Air Blank	0.000	10:20:57
Subject Test	0.169*	10:24:04
Air Blank	0.000	10:24:37
Air Blank	0.000	10:25:04
Subject Test	XXX**	10:25:24
Air Blank	RFI***	10:25:31

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. PLODYETT III AGENCY LHCPD
DATE 01-21-19 TIME 1027
INTOXILYZER SERIAL # 80-003856 LOCATION LHCPD JAIL

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0.100 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

Historical Note
New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

TRIMMZYER 8000

Location: JAIL

Serial Number: 87-03355

Print Version: A1.05.48

01/01/2019

10:27:36

Standard Lot#: 15917104A7

Last Changed By: J. SONDROE, #187

QAS: L. BUNKETT III, #253

LAKE HAVASU CITY PD

Operator: L. BUNKETT III, #253

LAKE HAVASU CITY PD

Test	g/100	Time
Air Blank	0.000	10:28:00
1.11.160 Blank	0.100	10:28:27
Air Blank	0.100	10:28:56

Preventative Maintenance
Performed by: Timmy Frazier