




POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Larry Allen is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

 #251
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 26 day of JANUARY.

2022 by LARRY ALLEN


NOTARY PUBLIC

1-26-2022
DATE

Commission expires:



Area Code 928

- | | | | | |
|-----------------------|----------|-----------------------------|-------------------------------|----------|
| Administration | 855-4884 | Fax #'s | Emergency | 9-1-1 |
| Business Office | 855-1171 | Administration | Non-Emergency Dispatch/ | 855-4111 |
| Investigations | 855-5775 | Business Office | Havasu Silent Witness | 854-TIPS |
| Patrol | 855-0515 | Investigations/Patrol | | |

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LARRY ALLEN AGENCY LWCPD
DATE 01-24-2022 TIME 1727
INTOXILYZER SERIAL # 80-003355 LOCATION LWCPD JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
01/24/2022 17:27:42

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: L. ALLEN #251
LAKE HAVASU CITY PD

Operator: L. ALLEN #251
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	17:28:25
0.100 Cal Check	0.101	17:28:46
Air Blank	0.000	17:29:15

Preventative Maintenance
Performed -- Timer Reset