



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy #281
 Signature

STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)

Subscribed and sworn to before me this 28th day of JANUARY

2022 by MICHAEL MALOY

Shawna Mcknight
 NOTARY PUBLIC

JANUARY 28, 2022
 DATE

Commission expires: MAY 4, 2025



Administration	855-4884	Area Code 928	Emergency	9-1-1
Business Office	855-1171	Fax #'s	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Administration	Havasu Silent Witness	854-TIPS
Patrol	855-0515	Business Office		
		Investigations/Patrol		

E-mail: police@lhcaz.gov

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 01-28-2022 TIME 2333
INTOXILYZER SERIAL # 90-003680

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)

Deficient Sample printed

2. Alcohol-free Subject Test (Proper Sample Recognition):

0. 000 AC

3. Mouth Alcohol Subject Test (Proper Sample Recognition):

Invalid Sample – Begin new deprivation period printed

4. Radio Frequency Interference Test (Error Recognition):

RFI Detect printed

5. Standard Calibration Check:

0. 100 AC

6. Air Blanks Completed

7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

PLACED AT SIGN B

SIGNATURE Michael Maloy #281

ARIZONA FORENSIC BREATH ALCOHOL
ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003680
Software Version: 9439.01.00
Analytical Report Number: 36800128220006
OAS: MALLOY, MICHAEL
OAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 01/06/2022
Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Test Date: 01/28/2022
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 23:15
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALLOY, MICHAEL
Agency: LAKE HAVASU PD
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 23621100A4
Expiration Date: 11/05/2023
Bottle #: 002
Last Changed By:
SLOWINSKI, JOSEPH
Permit #: 10002

Test	g/210L	Time
Air Blank	0.000	23:33:15
Diagnostic Test	PASS	23:33:51
Air Blank	0.000	23:34:28
Calibration Chk	0.100	23:34:50
Air Blank	0.000	23:35:29
Subject Test 1	DEF*	23:38:44
Air Blank	0.000	23:39:33
Air Blank	0.000	23:40:13
Subject Test 2	0.000	23:40:51
Air Blank	0.000	23:41:34
Wait		23:45:37
Air Blank	0.000	23:46:15
Subject Test 3	INV**	23:46:53
Air Blank	RFI***	23:47:25

*Deficient Sample

**Invalid Sample - Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ
by more than 0.020 g/210L.
Standard check results must be ±10%.