



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.


Michael T. Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 10 day of JULY
2024 by MICHAEL MALOY

[Signature]
NOTARY PUBLIC

7-10-2024
DATE



Commission expires: 3-16-2025

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

E-mail: police@lhcaz.gov

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPS JAIL
DATE 07-10-2024 TIME 19:10
INTOXILYZER SERIAL # 90-003679

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0.038 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

P.S.I: 625

SIGNATURE

Michael Maloy Jr #281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
 Serial Number: 90-003679
 Software Version: 9439.01.00
 Analytical Report Number: 36790710240275

QAS: MALOY, MICHAEL
 QAS Permit #: 30178
 Agency: LAKE HAVASU PD
 Last 31-Day Check: 06/10/2024
 Last Annual Maintenance: 09/22/2023

SUBJECT INFORMATION

Name: 31-Day Check
 Test Date: 07/10/2024
 Driver's License #:
 Gender:
 Date of Birth:
 Age:

Weight:
 State of Issue:
 Driver's License Expiration:
 Deprivation Start Time: 18:45
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
 Agency: LAKE HAVASU PD

Permit #: 30178
STANDARD INFORMATION
 Standard Value: 0.100
 Standard Lot #: 302-402720275
 Expiration Date: 05/24/2025
 Bottle #: 003
 Last Changed By: MALOY, MICHAEL

Test	g/210L	Time
Air Blank	0.000	19:10:32
Diagnostic Test	PASS	19:11:07
Air Blank	0.000	19:11:44
Calibration Chk	0.098	19:12:06
Air Blank	0.000	19:12:46
Subject Test 1	DEF*	19:16:02
Air Blank	0.000	19:16:43
Air Blank	0.000	19:17:24
Subject Test 2	0.000	19:17:59
Air Blank	0.000	19:18:42
Wait		19:22:46
Air Blank	0.000	19:23:23
Subject Test 3	INV**	19:24:02
Air Blank	RFI***	19:24:15

*Deficient Sample
 **Invalid Sample - Begin new deprivation period
 ***RFI Detect

RESULTSEXCEPTION MESSAGES

* Deficient Sample
 ** Invalid Sample - Begin new deprivation period
 *** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
 Consecutive subject test results must not differ by more than 0.020 g/210L.
 Standard check results must be $\pm 10\%$.*