



POLICE DEPARTMENT
LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature


STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 17th day of July.

2021 by MICHAEL MALOY

[Signature]
NOTARY PUBLIC

7/17/2021
DATE



Commission expires: 3/16/2025

Area Code 928

Administration 855-4884	Fax #'s	Emergency 9-1-1
Business Office 855-1171	Administration 680-5430	Non-Emergency Dispatch/ 855-4111
Investigations 855-5775	Business Office 680-5431	Havasu Silent Witness 854-TIPS
Patrol 855-0515	Investigations/Patrol 680-5432	

E-mail: police@lhcaz.gov

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO
DATE 07-17-2021 TIME 1520
INTOXILYZER SERIAL # 80-003355 LOCATION LHCPO JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
07/17/2021 15:20:46

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:21:17
0.100 Cal Check	0.100	15:21:37
Air Blank	0.000	15:22:07

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO

DATE 07-17-2021 TIME 1523

INTOXILYZER SERIAL # 80-003355 LOCATION LHCPO JAIL

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
07/17/2021

15:23:44

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:24:38
Diagnostic Test	Pass	15:25:13
Air Blank	0.000	15:25:40
0.100 Cal Check	0.100	15:26:01
Air Blank	0.000	15:26:31
Subject Test	0.000	15:26:52
Air Blank	ABT*	15:26:59
Air Blank	IPS**	15:27:07

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
 Location: JAIL
 Serial Number: 80-003355
 Core Version: 8105.48
 07/17/2021

15:27:53

Standard Lot#: 35019100A1
 Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
 LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
 LAKE HAVASU CITY PD

Subject: SQAP,TEST

DOB: 01/01/2000

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:28:40
Diagnostic Test	Pass	15:29:15
Air Blank	0.000	15:29:43
0.100 Cal Check	0.100	15:30:04
Air Blank	0.000	15:30:33
Subject Test	0.000*	15:33:41
Air Blank	0.000	15:34:08
Air Blank	0.000	15:34:36
Subject Test	XXX**	15:34:51
Air Blank	RFI***	15:35:03

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LKCPD

DATE 07-17-2021 TIME 1535

INTOXILYZER SERIAL # 80-003355 LOCATION LKCPD JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000

Location: JAIL

Serial Number: 80-003355

Core Version: 8105.48

07/17/2021

15:35:39

Standard Lot#: 35019100A1

Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281

LAKE HAVASU CITY PD

Operator: M. MALOY JR #281

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:36:11
0.100 Cal Check	0.099	15:36:31
Air Blank	0.000	15:37:01

Preventative Maintenance
Performed -- Timer Reset