



POLICE DEPARTMENT  
**LAKE HAVASU CITY**  
2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281  
Signature

STATE OF ARIZONA        )  
  ) ss  
COUNTY OF MOHAVE     )


Subscribed and sworn to before me this 30th day of July.

2022 by MICHAEL MALOY

Shawna MCKNIGHT  
NOTARY PUBLIC

July 30, 2022  
DATE

Commission expires: MAY 4, 2025



Area Code 928

Administration .....	855-4884	Fax #'s		Emergency .....	9-1-1
Business Office .....	855-1171	Administration .....	680-5430	Non-Emergency Dispatch/ .....	855-4111
Investigations .....	855-5775	Business Office .....	680-5431	Havasu Silent Witness .....	854-TIPS
Patrol .....	855-0515	Investigations/Patrol .....	680-5432		

E-mail: police@lhcaz.gov

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPS  
DATE 07-27-2022 TIME 0849  
INTOXILYZER SERIAL # 90-003680

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0. 0910 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Michael Maloy Jr

281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003680  
Software Version: 9439.01.00  
Analytical Report Number: 36800727220021

QAS: MALOY, MICHAEL  
QAS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 06/30/2022  
Last Annual Maintenance: 11/15/2021

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 07/27/2022  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 08:30  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
Agency: LAKE HAVASU PD

Permit #: 30178

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 23621100A4  
Expiration Date: 11/05/2023  
Bottle #: 003  
Last Changed By: MALOY, MICHAEL

Permit #: 30178

## RESULTS

Test	g/210L	Time
Air Blank	0.000	08:49:29
Diagnostic Test	PASS	08:50:05
Air Blank	0.000	08:50:42
Calibration Chk	0.096	08:51:03
Air Blank	0.000	08:51:43
Subject Test 1	DEF*	08:54:55
Air Blank	0.000	08:55:37
Air Blank	0.000	08:56:18
Subject Test 2	0.000	08:56:53
Air Blank	0.000	08:57:35
Wait		09:01:38
Air Blank	0.000	09:02:15
Subject Test 3	INV**	09:02:55
Air Blank	RFI***	09:03:10

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*