



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Josh Sondrol
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 24 day of July.

2019 by Josh Sondrol

Carla Duarte
NOTARY PUBLIC

July 24, 2019
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY ARIZONA ADMINISTRATIVE CODE
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Sanchez 187 AGENCY LUC PD

DATE 7/3/10 TIME 1932

INTOXILYZER SERIAL # 30-001289 LOCATION Lucasville

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .102 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 00 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18

gs187

Department of Public Safety - Alcohol Testing

Exhibit G-4. Standard Quality Assurance Procedure Intoxilyzer, Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Sondrol #187 AGENCY LHC PD
DATE 7/3/19 TIME 1930
INTOXILYZER SERIAL # 80100223 LOCATION JAIL

INTOXILYZER 8000
Location: CHANNEL BOOKING
Serial Number: 80-001029
Core Version: 8135.48
07/03/2019 10:34:11

(X) 1. Display Reads "PUSH BUTTON TO START"

Standard Lot#: 13518100A1
Last Changed By: L. PLUNKETT III #253

DIAGNOSTIC TESTS

- (X) 1. Clock time check.
- (X) 2. Date check.

DPS: J. SONDROL #187
LAKE HAVASU CITY PD

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0. 000 AC.
- (X) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (X) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (X) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 093 AC.

Operator: J. SONDROL #187
LHC PD

Subject: SOAP_TEST
DOB: 01/01/1950
Sex: M Weight: 123

Instrument is operating properly and accurately. YES NO

15 Minute Deprivation Period? Yes

COMMENTS

SIGNATURE J. Sondrol

DPS Form Exh G-4 (Rev 05-01)

Test	g/210L	Time
Air Blank	0.000	10:35:12
Diagnostic Test	Pass	10:35:49
Air Blank	0.000	10:36:16
0.100 Cal Check	0.098	10:36:37
Air Blank	0.000	10:37:07
Subject Test	0.080*	10:40:14
Air Blank	0.000	10:40:46
Air Blank	0.000	10:41:13
Subject Test	0.078	10:41:39
Air Blank	0.000	10:42:12
Wait		
Air Blank	0.000	10:47:06
Subject Test	XXX**	10:47:26
Air Blank	RF:***	10:47:34

Historical Note

New Exhibit G-4 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2011

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

INTOXILYZER 8000
Location: CHANNEL BOOKING
Serial Number: 80-001029
Core Version: 8135.48
07/03/2019 10:33:14

Standard Lot#: 13518100A1
Last Changed By: L. PLUNKETT III #253

DPS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SOAP_TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period: Yes

Test	g/210L	Time
Air Blank	0.000	10:30:58
Diagnostic Test	Pass	10:31:36
Air Blank	0.000	10:32:03
0.100 Cal Check	0.098	10:32:24
Air Blank	0.000	10:32:54
Subject Test	0.000	10:33:11
Air Blank	RF:*	10:33:18
Air Blank	IPS**	10:33:25

*Sequence Aborted
**Improper Sample

Not a Successfully
Completed
Test Sequence

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R.13-20-100(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. SONDRA 187 AGENCY LS2 PD

DATE 11-13-19 TIME 1048

INTOXILYZER SERIAL # 36-00029 LOCATION INR002

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR

Pour a standard alcohol concentration solution 100 AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., Me

[Signature]