



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael T. Maloy Jr #2BI
Signature


STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 12 day of JUNE.

2025 by MICHAEL MALOY
Jennifer Lentz
NOTARY PUBLIC

06/12/2025
DATE

Commission expires: 03/15/2029



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

E-mail: police@lhcaz.gov

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasut Silent Witness 854-TIPS

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY UNCPLD
DATE 06-12-2025 TIME 18:43
INTOXILYZER SERIAL # 90-003696

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0.101 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS: STANDARD CHANGE PSI 850

SIGNATURE Michael Maloy #281
DPS Form Exh I-2 (Rev 19-03)

zel

ARIZONA FORENSIC BREATH ALCOHOL
ANALYTICAL REPORT

STANDARD
CHANGE

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-00369G

Test Date: 06/12/2025

Test Time: 18:40

Location:

Software Version: 9439.01.00

Standard Lot #: 302-403183349

Standard Value: 0.100

Expiration Date: 01/24/2027

Bottle #: 048

Last Changed By: MALOY, MICHAEL

Permit #: 30178

Operator: MALOY, MICHAEL

Permit #:30178

Agency: LAKE HAVASU PD

Test	g/210L	Time
Air Blank	0.000	18:40:35
Calibration Chk	0.101	18:40:57
Air Blank	0.000	18:41:37

No RFI Detected

Successfully Completed Test Sequence

INSTRUMENT INFORMATION
Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-00369G
Software Version: 9439.01.00
Analytical Report Number: 36960612250081
OAS: MALOY, MICHAEL
OAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 12/03/2024
Last Annual Maintenance: 12/09/2024

SUBJECT INFORMATION

Test Date: 06/12/2025

Name: 31-Day Check,

Driver's License #:

Weight:

Gender:

State of Issue:

Date of Birth:

Driver's License Expiration:

Age:

Deprivation Start Time: 18:15

15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL

Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100

Standard Lot #: 302-403183349

Expiration Date: 01/24/2027

Bottle #: 048

Last Changed By:

MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	18:43:51
Diagnostic Test	PASS	18:44:27
Air Blank	0.000	18:45:04
Calibration Chk	0.101	18:45:25
Air Blank	0.000	18:46:06
Subject Test 1	DEF*	18:50:26
Air Blank	0.000	18:51:11
Air Blank	0.000	18:51:51
Subject Test 2	0.000	18:52:38
Air Blank	0.000	18:53:19
Wait		18:57:23
Air Blank	0.000	18:58:00
Subject Test 3	INV**	18:59:29
Air Blank	RFI***	18:59:42

*Deficient Sample

**Invalid Sample - Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample

** Invalid Sample - Begin new deprivation period

*** RFI Detect

All air blank results must be 0.000.

Consecutive subject test results must not differ
by more than 0.020 g/210L.

Standard check results must be $\pm 10\%$.