



POLICE DEPARTMENT

LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 17th day of JUNE.

2023 by MICHAEL MALOY

Jennifer Le Lentz
NOTARY PUBLIC

6/17/2023
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

E-mail: police@lhcaz.gov

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LMCPD
DATE 06-17-2023 TIME 0715
INTOXILYZER SERIAL # 90-003635

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0.100 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE

Michael Maldy Jr #281

281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003695
Software Version: 9439.01.00
Analytical Report Number: 36950617230146

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 05/24/2023
Last Annual Maintenance: 11/28/2022

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/17/2023
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 06:55
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402454776
Expiration Date: 06/16/2024
Bottle #: 001
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	07:15:55
Diagnostic Test	PASS	07:16:31
Air Blank	0.000	07:17:08
Calibration Chk	0.100	07:17:30
Air Blank	0.000	07:18:09
Subject Test 1	DEF*	07:21:25
Air Blank	0.000	07:22:11
Air Blank	0.000	07:22:52
Subject Test 2	0.000	07:23:16
Air Blank	0.000	07:23:57
Wait		07:28:00
Air Blank	0.000	07:28:37
Subject Test 3	INV**	07:29:15
Air Blank	RFI***	07:29:28

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTSEXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.