



Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY UPLD
DATE 06-17-2023 TIME 0636
INTOXILYZER SERIAL # 90-003816

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

☒ 1. Deficient Subject Test (Proper Sample Recognition)

Deficient Sample printed

☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC

☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):

Invalid Sample – Begin new deprivation period printed

☒ 4. Radio Frequency Interference Test (Error Recognition):

RFI Detect printed

☒ 5. Standard Calibration Check:
0.101 AC

☒ 6. Air Blanks Completed

☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE Michael Maloy Jr #281
DPS Form Exh I-2 (Rev 19-03)

INSTRUMENT INFORMATION

Analytical Instrument: INTOXLYZER 9000
Serial Number: 90-003696
Software Version: 9439.01.00
Analytical Report Number: 36960617230006
QAS: MALOV, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 05/19/2023
Last Annual Maintenance: 12/08/2022

SUBJECT INFORMATION

Test Date: 06/17/2023
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 06:15
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOV, MICHAEL
Agency: LAKE HAVASU PD
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402454776
Expiration Date: 06/16/2024
Bottle #: 007
Last Changed By:
MALOV, MICHAEL
Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	06:36:18
Diagnostic Test	PASS	06:36:54
Air Blank	0.000	06:37:31
Calibration Chk	0.101	06:37:52
Air Blank	0.000	06:38:32
Subject Test 1	DEF*	06:41:44
Air Blank	0.000	06:42:25
Air Blank	0.000	06:43:06
Subject Test 2	0.000	06:43:29
Air Blank	0.000	06:44:13
Wait		06:48:16
Air Blank	0.000	06:48:53
Subject Test 3	INV**	06:50:26
Air Blank	RFI***	06:50:41

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ
by more than 0.020 g/210L.
Standard check results must be ±10%.