



Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. BLUNKETT III AGENCY LHCPD

DATE 06-02-18 TIME 1158

INTOXILYZER SERIAL # 80-003355 LOCATION LHCPD JAIL

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

(X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"


(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(X) 4. Air blank completed.

(X) 5. Calibration check completed. Test results 0. 100 AC.

(X) 6. Air blank completed.

(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000  
Location: JAIL  
Serial Number: 80-003355  
Core Version: 8105.48  
06/02/2018

11:58:30

-----  
Standard Lot#: 13917100A1  
Last Changed By: D. SPIVEY #T17  
-----

QAS: L. PLUNKETT III #253  
LAKE HAVASU CITY PD  
-----

Operator: L. PLUNKETT III #253  
LAKE HAVASU CITY PD  
-----

Test	g/210L	Time
Air Blank	0.000	11:59:02
0.100 Cal Check	0.100	11:59:22
Air Blank	0.000	11:59:52

Preventative Maintenance  
Performed -- Timer Reset