



POLICE DEPARTMENT

LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael T. Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 3rd day of JUNE.

2025 by MICHAEL MALOY

Jennifer Lentz
NOTARY PUBLIC

06/03/25
DATE



Commission expires: 03/15/2029

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

E-mail: police@lhcaz.gov

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasus Silent Witness 854-TIPS

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LHCPD
DATE 06-02-2025 TIME 15:15
INTOXILYZER SERIAL # 90-003679

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.100 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS: INSTRUMENT TAKEN OUT OF SERVICE P.S.I.: 900
FOR END OF MEMORIAL DAY WEEKEND

SIGNATURE

Michael Maldy #281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
 Serial Number: 90-003679
 Software Version: 9439.01.00
 Analytical Report Number: 36790602250297

QAS: MALOY, MICHAEL
 QAS Permit #: 30178
 Agency: LAKE HAVASU PD
 Last 31-Day Check: 05/19/2025
 Last Annual Maintenance: 08/14/2024

SUBJECT INFORMATION

Name: 31-Day Check
 Test Date: 06/02/2025
 Driver's License #:
 Gender:
 Date of Birth:
 Age:

Weight:
 State of Issue:
 Driver's License Expiration:
 Deprivation Start Time: 15:00
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
 Standard Lot #: 302-403183349
 Expiration Date: 01/24/2027
 Bottle #: 048
 Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	15:15:37
Diagnostic Test	PASS	15:16:12
Air Blank	0.000	15:16:49
Calibration Chk	0.100	15:17:11
Air Blank	0.000	15:17:52
Subject Test 1	DEF*	15:21:54
Air Blank	0.000	15:22:51
Air Blank	0.000	15:23:31
Subject Test 2	0.000	15:25:00
Air Blank	0.000	15:25:46
Wait		15:29:51
Air Blank	0.000	15:30:28
Subject Test 3	INV**	15:31:02
Air Blank	RFI***	15:31:15

*Deficient Sample
 **Invalid Sample - Begin new deprivation period
 ***RFI Detect

RESULTSEXCEPTION MESSAGES

* Deficient Sample
 ** Invalid Sample - Begin new deprivation period
 *** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
 Consecutive subject test results must not differ by more than 0.020 g/210L.
 Standard check results must be $\pm 10\%$.