

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LUCPD

DATE 06-30-2021 TIME 1626 HRS

INTOXILYZER SERIAL # 80-001029 LOCATION CHANNEL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy Jr #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001029
Core Version: 8105.48
05/30/2021 16:26:03

Standard Lot#: 35019100A1
Last Changed By: K. TREBES #7259

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	16:26:52
0.100 Cal Check	0.100	16:27:13
Air Blank	0.000	16:27:42

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 06-30-2021 TIME 1635
INTOXILYZER SERIAL # 80-001029 LOCATION CHANNEL

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Michael Maloy Jr #281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001029
Core Version: 8105.48
06/30/2021 16:35:14

Standard Lot#: 35019100A1
Last Changed By: K. TREBES #7259

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SOAP, TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:36:12
Diagnostic Test	Pass	16:36:50
Air Blank	0.000	16:37:17
0.100 Cal Check	0.098	16:37:38
Air Blank	0.000	16:38:07
Subject Test	0.000	16:38:26
Air Blank	ABT*	16:38:32
Air Blank	IPS**	16:38:40

*Sequence Aborted
**Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001029
Core Version: 8105.48
06/30/2021 16:39:04

Standard Lot#: 35019100A1
Last Changed By: K. TREBES #7259

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SOAP, TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:39:54
Diagnostic Test	Pass	16:40:31
Air Blank	0.000	16:40:59
0.100 Cal Check	0.098	16:41:20
Air Blank	0.000	16:41:49
Subject Test	0.000*	16:44:57
Air Blank	0.000	16:45:24
Air Blank	0.000	16:45:51
Subject Test	XXX**	16:46:19
Air Blank	RFI***	16:46:30

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully
Completed
Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO

DATE 06-30-2021 TIME 1647

INTOXILYZER SERIAL # 80-001029 LOCATION CHANNEL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy Jr #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001029
Core Version: 8105.48
06/30/2021 16:47:02

Standard Lot#: 35019100A1
Last Changed By: K. TREBES #7259

OAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	16:47:31
0.100 Cal Check	0.098	16:47:52
Air Blank	0.000	16:48:22

Preventative Maintenance
Performed -- Timer Reset