



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
 Signature

STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)


Subscribed and sworn to before me this 12 day of MARCH.

2026 by MICHAEL MALOY

Jennifer Lentz
 NOTARY PUBLIC

03/12/2026
 DATE

Commission expires:
03/15/2029



Area Code 928

Administration	855-4884	Fax #'s	Emergency	9-1-1
Business Office	855-1171	Administration	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	Havasu Silent Witness	854-TIPS
Patrol	855-0515	Investigations/Patrol		

E-mail: police@lhcaz.gov

281

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO
DATE 03-12-2020 TIME 20:56
INTOXILYZER SERIAL # 90-003695

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS: P.S.I. 900

SIGNATURE Michael Maloy #281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
 Serial Number: 90-003695
 Software Version: 9439.01.00
 Analytical Report Number: 36950312260433

QAS: MALOY, MICHAEL
 QAS Permit #: 30178
 Agency: LAKE HAVASU PD
 Last 31-Day Check: 02/15/2026
 Last Annual Maintenance: 10/08/2025

SUBJECT INFORMATION

Name: 31-Day Check
 Test Date: 03/12/2026
 Driver's License #:
 Gender:
 Date of Birth:
 Age:

Weight:
 State of Issue:
 Driver's License Expiration:
 Deprivation Start Time: 20:30
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
 Standard Lot #: 302-403434231
 Expiration Date: 10/03/2027
 Bottle #: 040
 Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	20:56:49
Diagnostic Test	PASS	20:57:25
Air Blank	0.000	20:58:02
Calibration Chk	0.098	20:58:23
Air Blank	0.000	20:59:03
Subject Test 1	DEF*	21:02:46
Air Blank	0.000	21:03:27
Air Blank	0.000	21:04:08
Subject Test 2	0.000	21:05:24
Air Blank	0.000	21:06:06
Wait		21:10:10
Air Blank	0.000	21:10:47
Subject Test 3	INV**	21:11:30
Air Blank	RFI***	21:11:43

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
 ** Invalid Sample - Begin new deprivation period
 *** RFI Detect

OPERATOR COMMENTS

*Deficient Sample
 **Invalid Sample - Begin new deprivation period
 ***RFI Detect

*All air blank results must be 0.000.
 Consecutive subject test results must not differ by more than 0.020 g/210L.
 Standard check results must be ±10%.*

Not a Successfully Completed Test Sequence