



POLICE DEPARTMENT  
**LAKE HAVASU CITY**  
2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:  
  
I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281  
Signature

STATE OF ARIZONA )  
 ) ss  
COUNTY OF MOHAVE )

Subscribed and sworn to before me this 18 day of MARCH.

2022 by MICHAEL MALOY

Carla Duarte  
NOTARY PUBLIC

DATE 3-18-22

Commission expires:



Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasu Silent Witness ..... 854-TIPS

E-mail: [police@lncaz.gov](mailto:police@lncaz.gov)

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD  
DATE 03-17-2022 TIME 2348  
INTOXILYZER SERIAL # 90-003695

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0. 101 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

CHANGED O2Y GAS STANDARD

SIGNATURE

Michael Maloy Jr #281

ARIZONA FORENSIC BREATH ALCOHOL  
ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000  
Serial Number: 90-003695  
Software Version: 9439.01.00  
Analytical Report Number: 36950317220004  
ORS: MRLDY, MICHAEL  
ORS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 12/10/2021  
Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Test Date: 03/17/2022  
Name: 31-Day Check,  
Driver's License #:  
Weight:  
Gender:  
State of Issue:  
Date of Birth:  
Driver's License Expiration:  
Age:  
Deprivation Start Time: 23:30  
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MRLDY, MICHAEL  
Agency: LAKE HAVASU PD  
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 1463724  
Expiration Date: 03/10/2023  
Bottle #: 003  
Last Changed By:  
MRLDY, MICHAEL  
Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	23:48:44
Diagnostic Test	PASS	23:49:20
Air Blank	0.000	23:49:56
Calibration Chk	0.101	23:50:18
Air Blank	0.000	23:50:58
Subject Test 1	DEF*	23:54:24
Air Blank	0.000	23:55:31
Air Blank	0.000	23:56:11
Subject Test 2	0.000	23:56:45
Air Blank	0.000	23:57:34
Wait		00:01:38
Air Blank	0.000	00:02:15
Subject Test 3	INV**	00:03:04
Air Blank	RF1***	00:03:22

\*Deficient Sample

\*\*Invalid Sample - Begin new deprivation period

\*\*\*RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

\* Deficient Sample

\*\* Invalid Sample - Begin new deprivation period

\*\*\* RFI Detect

All air blank results must be 0.000.

Consecutive subject test results must not differ  
by more than 0.020 g/210L.

Standard check results must be ±10%.

Intoxilyzer - Alcohol Analyzer  
Model 9000

SN: 90-003695  
Test Date: 03/17/2022  
Test Time: 23:43

Location:

Software Version: 9439.01.00

Standard Lot #: 1463724  
Standard Value: 0.100  
Expiration Date: 03/10/2023  
Bottle #: 003  
Last Changed By: MRLDY, MICHAEL  
Permit #: 30178

Operator: MRLDY, MICHAEL  
Permit #: 30178  
Agency: LAKE HAVASU PD

Test	g/210L	Time
Air Blank	0.000	23:44:18
Calibration Chk	0.102	23:44:40
Air Blank	0.000	23:45:19

No RFI Detected

Successfully Completed Test Sequence