



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy #281
 Signature

STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)

Subscribed and sworn to before me this 23 day of MARCH.

2022 by MICHAEL MALOY

Carla Duarte
 NOTARY PUBLIC

3-23-22
 DATE

Commission expires:



Area Code 928

- | | | | | |
|-----------------------|----------|-----------------------------|-------------------------------|----------|
| Administration | 855-4884 | Fax #'s | Emergency | 9-1-1 |
| Business Office | 855-1171 | Administration | Non-Emergency Dispatch/ | 855-4111 |
| Investigations | 855-5775 | Business Office | Havasu Silent Witness | 854-TIPS |
| Patrol | 855-0515 | Investigations/Patrol | | |

E-mail: police@lhcaz.gov

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 03-23-2022 TIME 1533
INTOXILYZER SERIAL # 90-003679

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.099 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE Michael Maloy #281

281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003679
Software Version: 9439.01.00
Analytical Report Number: 36790323220023

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 02/23/2022
Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 03/23/2022
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 15:15
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 23621100A4
Expiration Date: 11/05/2023
Bottle #: 003
Last Changed By: SLOWINSKI, JOSEPH

Permit #: 10002

Test	g/210L	Time
Air Blank	0.000	15:33:02
Diagnostic Test	PASS	15:33:39
Air Blank	0.000	15:34:16
Calibration Chk	0.099	15:34:37
Air Blank	0.000	15:35:17
Subject Test 1	DEF*	15:38:36
Air Blank	0.000	15:39:23
Air Blank	0.000	15:40:04
Subject Test 2	0.000	15:40:32
Air Blank	0.000	15:41:13
Wait		15:45:17
Air Blank	0.000	15:45:54
Subject Test 3	INV**	15:46:24
Air Blank	RFI***	15:46:36

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.

Not a Successfully Completed Test Sequence