



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 1 day of MAY

2020 by MICHAEL MALOY

Jennifer Lentz
NOTARY PUBLIC

05.01.2020
DATE



Commission expires: 03.15.2029

Area Code 928

- | | | | | |
|-----------------------|----------|-----------------------------|-------------------------------|----------|
| Administration | 855-4884 | Fax #'s | Emergency | 9-1-1 |
| Business Office | 855-1171 | Administration | Non-Emergency Dispatch/ | 855-4111 |
| Investigations | 855-5775 | Business Office | Havasu Silent Witness | 854-TIPS |
| Patrol | 855-0515 | Investigations/Patrol | | |

E-mail: police@lhcaz.gov

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 05-01-2026 TIME 01:32
INTOXILYZER SERIAL # 90-003695

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 096 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

P.S.I.: 800

SIGNATURE Michael Maloy Jr #281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003695
Software Version: 9439.01.00
Analytical Report Number: 36950501260460

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 04/09/2026
Last Annual Maintenance: 10/08/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 05/01/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 01:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 302-403434231
Expiration Date: 10/03/2027
Bottle #: 040
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	01:32:58
Diagnostic Test	PASS	01:33:34
Air Blank	0.000	01:34:11
Calibration Chk	0.096	01:34:32
Air Blank	0.000	01:35:12
Subject Test 1	DEF*	01:38:35
Air Blank	0.000	01:39:23
Air Blank	0.000	01:40:03
Subject Test 2	0.000	01:40:28
Air Blank	0.000	01:41:09
Wait		01:45:13
Air Blank	0.000	01:45:50
Subject Test 3	INV**	01:46:20
Air Blank	RFI***	01:46:33

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*