



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael T. Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 19 day of MAY.

2023 by MICHAEL MALOY
Carla Scoote
NOTARY PUBLIC

MAY 19, 2023
DATE

Commission expires:



Administration 855-4884 Fax #'s
Business Office 855-1171 Administration 680-5430
Investigations 855-5775 Business Office 680-5431
Patrol 855-0515 Investigations/Patrol 680-5432

Area Code 928

E-mail: police@lhcaz.gov

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY 281 AGENCY LHCPD
DATE 05-19-2023 TIME 0807
INTOXILYZER SERIAL # 90-003679

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.099 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE

Michael Maloy #281

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003679
Software Version: 9439.01.00
Analytical Report Number: 36790519230124
QAS: MALLOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 12/07/2022
Last Annual Maintenance: 11/29/2022

SUBJECT INFORMATION

Test Date: 05/19/2023
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 07:45
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALLOY, MICHAEL
Agency: LAKE HAVASU PD
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 1463724
Expiration Date: 06/10/2023
Bottle #: 004
Last Changed By:
MALLOY, MICHAEL
Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	08:07:39
Diagnostic Test	PASS	08:08:14
Air Blank	0.000	08:08:51
Calibration Chk	0.099	08:09:12
Air Blank	0.000	08:09:52
Subject Test 1	DEF*	08:13:54
Air Blank	0.000	08:14:39
Air Blank	0.000	08:15:19
Subject Test 2	0.000	08:15:42
Air Blank	0.000	08:16:22
Wait		08:20:26
Air Blank	0.000	08:21:03
Subject Test 3	INV**	08:21:26
Air Blank	RFI***	08:21:40

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ
by more than 0.020 g/210L.
Standard check results must be ±10%.