



E-mail: police@lhcaz.gov

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY 281 AGENCY LHCPD
DATE 05-19-2023 TIME 0823
INTOXILYZER SERIAL # 90-003696

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

☒ 1. Deficient Subject Test (Proper Sample Recognition)

Deficient Sample printed

☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC

☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):

Invalid Sample – Begin new deprivation period printed

☒ 4. Radio Frequency Interference Test (Error Recognition):

RFI Detect printed

☒ 5. Standard Calibration Check:
0.101 AC

☒ 6. Air Blanks Completed

☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

STANDARD CHANGE

SIGNATURE

Michael Maury #281

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003696
Software Version: 9439.01.00
Analytical Report Number: 36960519230005
OAS: MALLOY, MICHAEL
OAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 05/19/2022
Last Annual Maintenance: 12/08/2022

SUBJECT INFORMATION

Test Date: 05/19/2023
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 08:00
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALLOY, MICHAEL
Agency: LAKE HAVASU PD
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402454776
Expiration Date: 06/16/2024
Bottle #: 007
Last Changed By:
MALLOY, MICHAEL
Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	08:23:44
Diagnostic Test	PASS	08:24:20
Air Blank	0.000	08:24:57
Calibration Chk	0.101	08:25:18
Air Blank	0.000	08:25:58
Subject Test 1	DEF*	08:29:18
Air Blank	0.000	08:30:04
Air Blank	0.000	08:30:44
Subject Test 2	0.000	08:31:47
Air Blank	0.000	08:32:28
Wait		08:36:33
Air Blank	0.000	08:37:10
Subject Test 3	INV**	08:37:42
Air Blank	RFI***	08:37:55

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ
by more than 0.020 g/210L.
Standard check results must be ±10%.