



POLICE DEPARTMENT
LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 22 day of May.

2020 by Ashley Rae Bonds

[Signature]
NOTARY PUBLIC

5-22-2020
DATE

Commission expires: May 29, 2021



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LULAD

DATE 05-22-2020 TIME 0542

INTOXILYZER SERIAL # 80-001873 LOCATION EAST BANK CHANNEL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001873
Core Version: 8105.48
05/22/2020 05:42:29

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	05:43:32
0.100 Cal Check	0.100	05:43:53
Air Blank	0.000	05:44:23

Preventative Maintenance
Performed -- Timer Reset

281

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 05-22-2020 TIME 0546
INTOXILYZER SERIAL # 80-001873 LOCATION EAST BANK CHANNEL

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Michael Maloy Jr #281

INTOXILYZER 8000
 Location: LHCPD
 Serial Number: 80-001873
 Core Version: 8105.48
 05/22/2020 05:50:24

 Standard Lot#: 35019100A1
 Last Changed By: M. MALOY JR #281

 OAS: M. MALOY JR #281
 LAKE HAVASU CITY PD

 Operator: M. MALOY JR #281
 LAKE HAVASU CITY PD

 Subject: SQAP,TEST
 DOB: 01/01/2000
 Sex: M Weight: 200

 15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	05:51:16
Diagnostic Test	Pass	05:51:53
Air Blank	0.000	05:52:21
0.100 Cal Check	0.099	05:52:42
Air Blank	0.000	05:53:11
Subject Test	0.000*	05:56:19
Air Blank	0.000	05:56:46
Air Blank	0.000	05:57:13
Subject Test	XXX**	05:57:27
Air Blank	RFI***	05:57:35

*Deficient Sample
 **Invalid Sample
 Begin new deprivation period
 ***RFI Detect

Not a Successfully
 Completed
 Test Sequence

INTOXILYZER 8000
 Location: LHCPD
 Serial Number: 80-001873
 Core Version: 8105.48
 05/22/2020 05:46:30

 Standard Lot#: 35019100A1
 Last Changed By: M. MALOY JR #281

 OAS: M. MALOY JR #281
 LAKE HAVASU CITY PD

 Operator: M. MALOY JR #281
 LAKE HAVASU CITY PD

 Subject: SQAP,TEST
 DOB: 01/01/2000
 Sex: M Weight: 200

 15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	05:47:39
Diagnostic Test	Pass	05:48:17
Air Blank	0.000	05:48:44
0.100 Cal Check	0.099	05:49:05
Air Blank	0.000	05:49:35
Subject Test	0.000	05:49:52
Air Blank	ABT*	05:49:58
Air Blank	IPS**	05:50:05

*Sequence Aborted
 **Improper Sample

Not a Successfully
 Completed
 Test Sequence

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD

DATE 05-22-2020 TIME 0558

INTOXILYZER SERIAL # 80-001873 LOCATION EAST BANK CHANNEL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy Jr #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001873
Core Version: 8105.48
05/22/2020 05:58:02

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	05:58:33
0.100 Cal Check	0.098	05:58:54
Air Blank	0.000	05:59:23

Preventative Maintenance
Performed -- Timer Reset