



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Charles

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 4 day of June.


2019 by JOSH SONDROL

Carla Duarte
NOTARY PUBLIC

6-4-19

DATE

Commission expires:



Area Code 928

- | | | | | | |
|-----------------------|----------|-----------------------------|----------|-------------------------------|----------|
| Administration | 855-4884 | Fax #'s | | Emergency | 9-1-1 |
| Business Office | 855-1171 | Administration | 680-5430 | Non-Emergency Dispatch/ | 855-4111 |
| Investigations | 855-5775 | Business Office | 680-5431 | Havasu Silent Witness | 854-TIPS |
| Patrol | 855-0515 | Investigations/Patrol | 680-5432 | | |

E-mail: police@lhcaz.gov

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS RECORD IS PREPARED PURSUANT TO A DUTY IMPOSED BY A STATUTE OF THE STATE OF ARIZONA.

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROGRAMS
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J Sondrol 187 AGENCY LNC PD
DATE 5/24/19 TIME 1039
INTOXILYZER SERIAL # 80-001873 LOCATION JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 093 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 11, 2016

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-001873
Core version: 8105.48
05/17/2019 10:39:55

Standard Lot#: 13917100A1
Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:42:08
0.100 Cal Check	0.098	10:42:29
Air Blank	0.000	10:42:58

Preventative Maintenance
Performed -- Timer Reset

Exhibit G-4. Standard Quality Assurance Procedure Intoxilyzer, Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J Sondrol #187 AGENCY LMC PD
DATE 5/21/19 TIME 1045
INTOXILYZER SERIAL # 80-001873 LOCATION JAIL

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (X) 1. Clock time check.
(X) 2. Date check.

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0.000 AC.
(X) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(X) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(X) 4. Standard Calibration Check standard 0.100 AC. Result 0.098 AC.

Instrument is operating properly and accurately. YES [checked] NO

COMMENTS IN SUBJECTS DUE TO OTHER KNOWING REF: 67804

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-001873
Core Version: 8105.48
05/17/2019

[Signature]

10:45:20

Standard Lot#: 13917100A;
Last Changed By: J. SONDROL #187

OAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

SIGNATURE [Signature]

DPS Form Exh G-4 (Rev 05-01)

Historical Note

New Exhibit G-4 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18

Table with columns: Test, g/210L, Time. Rows include: Air Blank (0.000, 10:55:15), Diagnostic Test (Pass, 10:55:52), Air Blank (0.000, 10:56:23), 0.100 Cal Check (0.098, 10:56:41), Air Blank (0.000, 10:57:10), Subject Test (0.086*, 11:00:18), Air Blank (0.000, 11:00:50), Air Blank (0.000, 11:01:17), Subject Test (XXX**, 11:01:46), Air Blank (RFI***, 11:02:09).

[Signature]

10:54:27

Exhibit G-2 - Standard Calibration Check Procedure, Intoxilyzer Model 8000

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J Sondrol 187 AGENCY LNC PD
DATE 5/26/19 TIME 1103
INTOXILYZER SERIAL # 80-001873 LOCATION JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

Historical Note
New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 20

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-001873
Core version: 8105.48
05/17/2019

J Sondrol 187
11:03:30

Standard Lot#: 13917100A!
Last Changed By: J. SONDROL #187

CAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	11:03:59
0.100 Cal Check	0.097	11:04:20
Air Blank	0.000	11:04:49

Preventative Maintenance
Performed -- Timer Reset