



POLICE DEPARTMENT  
**LAKE HAVASU CITY**  
2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281  
Signature

STATE OF ARIZONA        )  
  ) ss  
COUNTY OF MOHAVE    )

Subscribed and sworn to before me this 2<sup>nd</sup> day of June.

2022 by Michael Maloy Jr.

[Signature]  
NOTARY PUBLIC

6-2-22  
DATE

Commission expires: 2-26-23



Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasu Silent Witness ..... 854-TIPS

E-mail: police@lhcaz.gov

## ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
 Serial Number: 90-003680  
 Software Version: 9439.01.00  
 Analytical Report Number: 36800529220016

QAS: MALOY, MICHAEL  
 QAS Permit #: 30178  
 Agency: LAKE HAVASU PD  
 Last 31-Day Check: 04/30/2022  
 Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Name: 31-Day Check  
 Test Date: 05/29/2022  
 Driver's License #:  
 Gender:  
 Date of Birth:  
 Age:

Weight:  
 State of Issue:  
 Driver's License Expiration:  
 Deprivation Start Time: 19:15  
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL  
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100  
 Standard Lot #: 23621100A4  
 Expiration Date: 11/05/2023  
 Bottle #: 003  
 Last Changed By: MALOY, MICHAEL

Permit #: 30178

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 19:35:38 |
| Diagnostic Test | PASS   | 19:36:13 |
| Air Blank       | 0.000  | 19:36:50 |
| Calibration Chk | 0.098  | 19:37:12 |
| Air Blank       | 0.000  | 19:37:51 |
| Subject Test 1  | DEF*   | 19:41:06 |
| Air Blank       | 0.000  | 19:41:48 |
| Air Blank       | 0.000  | 19:42:28 |
| Subject Test 2  | 0.000  | 19:42:54 |
| Air Blank       | 0.000  | 19:43:35 |
| Wait            |        | 19:47:39 |
| Air Blank       | 0.000  | 19:48:16 |
| Subject Test 3  | INV**  | 19:48:57 |
| Air Blank       | RFI*** | 19:49:13 |

RESULTSEXCEPTION MESSAGES

\*Deficient Sample  
 \*\*Invalid Sample - Begin new deprivation period  
 \*\*\*RFI Detect

\* Deficient Sample  
 \*\* Invalid Sample - Begin new deprivation period  
 \*\*\* RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.  
 Consecutive subject test results must not differ by more than 0.020 g/210L.  
 Standard check results must be  $\pm 10\%$ .

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LINCOPD  
DATE 05-29-2022 TIME 1935  
INTOXILYZER SERIAL # 90-003680

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0. 098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Michael Maloy Jr #281