



POLICE DEPARTMENT

LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr. #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 16th day of November 2023

_____ by Michael Maloy Jr.

[Signature]
NOTARY PUBLIC

11-16-23
DATE

Commission expires: 2/26/27



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

281

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALLOY #281 AGENCY LHCPD
DATE 11-16-2023 TIME 12:05
INTOXILYZER SERIAL # 90-003696

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. 098 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

750 P.S.I.

SIGNATURE

Michael Malloy #281

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003696
Software Version: 9439.01.00
Analytical Report Number: 36961116230020
QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 07/14/2023
Last Annual Maintenance: 12/08/2022

281

SUBJECT INFORMATION

Test Date: 11/16/2023
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 11:45
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402454776
Expiration Date: 06/16/2024
Bottle #: 007
Last Changed By:
MALOY, MICHAEL
Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	12:05:40
Diagnostic Test	PASS	12:06:16
Air Blank	0.000	12:06:53
Calibration Chk	0.098	12:07:14
Air Blank	0.000	12:07:54
Subject Test 1	DEF*	12:11:44
Air Blank	0.000	12:12:26
Air Blank	0.000	12:13:06
Subject Test 2	0.000	12:13:48
Air Blank	0.000	12:14:29
Wait		12:18:33
Air Blank	0.000	12:19:10
Subject Test 3	INV**	12:19:37
Air Blank	RFI***	12:19:51

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ