



POLICE DEPARTMENT

LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 23rd day of November.

2023 by Michael Maloy

Shawna McKnight
NOTARY PUBLIC

November 23rd 2023
DATE

Commission expires: May 4, 2025



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

281

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LNCPS
DATE 11-23-2023 TIME 10:20
INTOXILYZER SERIAL # 90-003680

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0.098 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

RETURNED FROM P.M.

P.S.I. 850 P.S.I

SIGNATURE M. MALOY #281 Michael Maloy #281

INSTRUMENT INFORMATION
Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003680
Software Version: 9439.01.00
Analytical Report Number: 36801123230034
QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 09/16/2023
Last Annual Maintenance: 11/06/2023

281

SUBJECT INFORMATION

Test Date: 11/23/2023
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 10:00
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD
Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402376471
Expiration Date: 6/2/2024
Bottle #: 053
Last Changed By:
MALOY, MICHAEL
Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	10:20:58
Diagnostic Test	PASS	10:21:33
Air Blank	0.000	10:22:10
Calibration Chk	0.098	10:22:32
Air Blank	0.000	10:23:11
Subject Test 1	DEF*	10:27:14
Air Blank	0.000	10:27:57
Air Blank	0.000	10:28:38
Subject Test 2	0.000	10:29:09
Air Blank	0.000	10:29:50
Wait		10:33:53
Air Blank	0.000	10:34:30
Subject Test 3	INV**	10:35:05
Air Blank	RFI***	10:35:17

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ