



ARIZONA DEPARTMENT OF PUBLIC SAFETY

INTOXILYZER MAINTENANCE REPORT

CR NO. I800033560006

Page 1 of 1

AGENCY Lake Havasu City Police Department
Lake Havasu City, AZ 86403
OFFICER PLUNKETT, #253
DATE October 09, 2018



INSTRUMENT

80003356

INSTRUMENT LOCATION

Lake Havasu City Police Department

MAINTENANCE REQUESTED

Preventative Maintenance

MAINTENANCE PERFORMED

Preventative Maintenance Performed:

Inspection, Calibration, Verification

A handwritten signature in black ink, appearing to read "Joseph M Slowinski".

JOSEPH M SLOWINSKI, #6101, Electronics Technician
Central Regional Crime Laboratory
2102 W. Encanto Blvd., Phoenix, AZ 85009
602-223-2159

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. PLUNKETT III AGENCY LHCPD

DATE 11-01-18 TIME 0721

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START" CONTACT GAS
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 253

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000
Location: MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
11/01/2018 07:21:05

Standard Lot#: 13917100A1
Last Changed By: J. SONDRON #187

OAS: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Operator: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	07:21:50
0.100 Cal Check	0.098	07:22:11
Air Blank	0.000	07:22:41

Preventative Maintenance
Performed -- Timer Reset

Exhibit G-4. Standard Quality Assurance Procedure Intoxilyzer, Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOUIS H. PHUNKET III AGENCY YHCPD

DATE 11-01-18 TIME 0730

INTOXILYZER SERIAL # 80-003356 LOCATION YHCPD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.

2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0.000 AC.

2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed

3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.

4. Standard Calibration Check standard 0.100 AC. Result 0.098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS RETURN FROM DPS ANNUAL PM/PREP FOR WET WORKSHOP AT
LONDON BRIDGE RESORT ON 11-02-18.

SIGNATURE  253

DPS Form Exh G-4 (Rev 05-01)

Historical Note

New Exhibit G-4 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000
Location: MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
11/01/2018 07:30:20

Standard Lot#: 13917100A1
Last Changed By: J. SONDROL #187

QAS: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Operator: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:31:54
Diagnostic Test	Pass	07:32:32
Air Blank	0.000	07:32:59
0.100 Cal Check	0.098	07:33:20
Air Blank	0.000	07:33:50
Subject Test	0.000	07:34:10
Air Blank	ABT*	07:34:17
Air Blank	IPS**	07:34:28

*Sequence Aborted
**Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
11/01/2018 07:37:29

Standard Lot#: 13917100A1
Last Changed By: J. SONDROL #187

QAS: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Operator: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:38:23
Diagnostic Test	Pass	07:39:00
Air Blank	0.000	07:39:28
0.100 Cal Check	0.098	07:39:48
Air Blank	0.000	07:40:18
Subject Test	0.065*	07:43:26
Air Blank	0.000	07:43:58
Air Blank	0.000	07:44:25
Subject Test	XXX**	07:44:43
Air Blank	RFI***	07:44:54

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully
Completed
Test Sequence

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. PLUNKETT III AGENCY LHCPD
DATE 11-01-18 TIME 0746
INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  253

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000
Location: MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
11/01/2018 07:46:23

Standard Lot#: 13917100A1
Last Changed By: J. SONDRON #187

QAS: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Operator: L. PLUNKETT III #253
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	07:46:54
0.100 Cal Check	0.098	07:47:15
Air Blank	0.000	07:47:45

**Preventative Maintenance
Performed -- Timer Reset**