



POLICE DEPARTMENT


**LAKE HAVASU CITY**

2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Larry Allen is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

  
\_\_\_\_\_  
Signature

STATE OF ARIZONA       )  
  ) ss  
COUNTY OF MOHAVE    )


Subscribed and sworn to before me this 18th day of November

2022 by LARRY ALLEN

Shawna McKnight  
NOTARY PUBLIC

11/18/2022  
DATE

Commission expires: May 4, 2025



Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasu Silent Witness ..... 854-TIPS

E-mail: [police@lhcaz.gov](mailto:police@lhcaz.gov)

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST L. ALLEN #251 AGENCY LAKE HAVASU CITY POLICE DEPT.  
DATE 10-26-2022 TIME 18:00  
INTOXILYZER SERIAL # 90-003695

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check  
☒ 2. Date check


OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed  
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC  
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed  
☒ 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed  
☒ 5. Standard Calibration Check:  
0.100 AC  
☒ 6. Air Blanks Completed  
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES X NO     

COMMENTS:

SIGNATURE  #251

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003695  
Software Version: 9439.01.00  
Analytical Report Number: 36951026220039

QAS: ALLEN, LARRY  
QAS Permit #: 30179  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 10/26/2022  
Last Annual Maintenance: 11/15/2021

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 10/26/2022  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 17:40  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: ALLEN, LARRY  
Agency: LAKE HAVASU PD

Permit #: 30179  
STANDARD INFORMATION  
Standard Value: 0.100  
Standard Lot #: 1463724  
Expiration Date: 03/10/2023  
Bottle #: 003  
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	18:00:41
Diagnostic Test	PASS	18:01:17
Air Blank	0.000	18:01:54
Calibration Chk	0.100	18:02:15
Air Blank	0.000	18:02:55
Subject Test 1	DEF*	18:06:18
Air Blank	0.000	18:07:35
Air Blank	0.000	18:08:15
Subject Test 2	0.000	18:08:44
Air Blank	0.000	18:09:26
Wait		18:13:31
Air Blank	0.000	18:14:08
Subject Test 3	INV**	18:14:29
Air Blank	RFI***	18:14:42

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

## RESULTS

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .*