



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
 Signature

STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)


Subscribed and sworn to before me this 15 day of September.

2021 by Ashley Rae Bonds

Ashley Rae Bonds
 NOTARY PUBLIC

9/15/21
 DATE

Commission expires: May 29, 2025



Area Code 928

Administration 855-4884	Fax #'s	Emergency 9-1-1
Business Office 855-1171	Administration 680-5430	Non-Emergency Dispatch/ 855-4111
Investigations 855-5775	Business Office 680-5431	Havasu Silent Witness 854-TIPS
Patrol 855-0515	Investigations/Patrol 680-5432	

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LHCPS
DATE 09-15-2021 TIME 1501
INTOXILYZER SERIAL # B0-003355 LOCATION LHCPS JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maldy #281

281

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
09/15/2021 15:01:56

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:02:30
0.100 Cal Check	0.100	15:02:51
Air Blank	0.000	15:03:21

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALLOY #281 AGENCY LHCPO
DATE 09-15-2021 TIME 1506
INTOXILYZER SERIAL # 80-003355 LOCATION LHCPO JAIL?

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 0.000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 095 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

TAKEN OUT OF SERVICE FOR ANNUAL
PREVENTATIVE MAINTENANCE

SIGNATURE

Michael Malloy #281

281

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
09/15/2021

15:06:16

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:07:19
Diagnostic Test	Pass	15:07:54
Air Blank	0.000	15:08:21
0.100 Cal Check	0.100	15:08:42
Air Blank	0.000	15:09:12
Subject Test	0.000	15:09:30
Air Blank	ABT*	15:09:39
Air Blank	IPS**	15:09:46

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

281

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
09/15/2021

15:10:27

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:11:18
Diagnostic Test	Pass	15:11:53
Air Blank	0.000	15:12:20
0.100 Cal Check	0.099	15:12:41
Air Blank	0.000	15:13:10
Subject Test	0.000*	15:16:18
Air Blank	0.000	15:16:45
Air Blank	0.000	15:17:13
Subject Test	XXX**	15:17:30
Air Blank	RFI***	15:17:44

*Deficient Sample
**Invalid Sample
 Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LUCPD
DATE 09-15-2025 TIME 1518
INTOXILYZER SERIAL # 81-003355 LOCATION LUCPD JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.095 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy Jr #281

INTOXILYZER 8000
Location: JAIL
Serial Number: 80-003355
Core Version: 8105.48
09/15/2021 15:18:19

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:18:49
0.100 Cal Check	0.095	15:19:10
Air Blank	0.000	15:19:40

Preventative Maintenance
Performed -- Timer Reset