

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPS

DATE 09-02-2020 TIME 2235

INTOXILYZER SERIAL # 80-001029 LOCATION CHANNEL MOBILE BOOKING

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001029
Core Version: 8105.48
09/02/2020 22:35:14

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

DAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	22:35:50
0.100 Ca: Check	0.097	22:36:10
Air Blank	0.000	22:36:40

Preventative Maintenance
Performed -- Timer Reset

281

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LHC PD

DATE 09-02-2020 TIME 2240

INTOXILYZER SERIAL # 80-001029 LOCATION CHANNEL MOBILE BOOKING

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001029
Core Version: 8105.48
09/02/2020 22:40:09

Standard Lot#: 35019100A!
Last Changed By: M. MALOY JR #281

OAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SOAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:41:10
Diagnostic Test	Pass	22:41:47
Air Blank	0.000	22:42:15
0.100 Cal Check	0.097	22:42:36
Air Blank	0.000	22:43:05
Subject Test	0.000	22:43:21
Air Blank	ABT*	22:43:28
Air Blank	!PS**	22:43:35

*Sequence Aborted
**Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: LHCPD
Serial Number: 80-001029
Core Version: 8105.48
09/02/2020 22:44:01

Standard Lot#: 35019100A!
Last Changed By: M. MALOY JR #281

OAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SOAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:44:45
Diagnostic Test	Pass	22:45:27
Air Blank	0.000	22:45:54
0.100 Cal Check	0.098	22:46:15
Air Blank	0.000	22:46:44
Subject Test	0.000*	22:49:52
Air Blank	0.000	22:50:19
Air Blank	0.000	22:50:46
Subject Test	XXX**	22:51:15
Air Blank	RFI***	22:51:29

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully
Completed
Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LUCPD

DATE 09-02-2020 TIME 2252

INTOXILYZER SERIAL # 80-601029 LOCATION CHANNEL MOBILE BOOKING

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maldy #281

281

INTOXILYZER 8000
Location: LHCPO
Serial Number: 80-001029
Core Version: 8105.48
09/02/2020 22:52:10

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

OAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test g/210L Time

Air Blank 0.000 22:52:40
0.100 Cal Check 0.097 22:53:01
Air Blank 0.000 22:53:31

Preventative Maintenance
Performed -- Timer Reset