

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LKCPD

DATE 09-02-2021 TIME 1702

INTOXILYZER SERIAL # 80-003356 LOCATION S16N B

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maldy #281

281

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
09/02/2021 17:02:00

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	17:02:31
0.100 Cal Check	0.098	17:02:52
Air Blank	0.000	17:03:21

Preventative Maintenance
Performed -- Timer Reset

281

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPO
DATE 09-02-2021 TIME 1704
INTOXILYZER SERIAL # 80-003356 LOCATION SLBN 8

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
09/02/2021

17:04:05

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:04:54
Diagnostic Test	Pass	17:05:29
Air Blank	0.000	17:05:56
0.100 Cal Check	0.099	17:06:17
Air Blank	0.000	17:06:47
Subject Test	0.000	17:07:05
Air Blank	ABT*	17:07:12
Air Blank	IPS**	17:07:19

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

281

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
09/02/2021

17:07:44

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:08:35
Diagnostic Test	Pass	17:09:10
Air Blank	0.000	17:09:37
0.100 Cal Check	0.098	17:09:58
Air Blank	0.000	17:10:27
Subject Test	0.000*	17:13:35
Air Blank	0.000	17:14:02
Air Blank	0.000	17:14:30
Subject Test	XXX**	17:14:43
Air Blank	RFI***	17:14:56

*Deficient Sample
**Invalid Sample
 Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

281

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALDY #281 AGENCY LVCPD
DATE 09-02-2021 TIME 1715
INTOXILYZER SERIAL # 80-003356 LOCATION SIGN B

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maldy #281

281

INTOXILYZER 8000
Location: CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
09/02/2021 17:15:38

Standard Lot#: 35019100A1
Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281
LAKE HAVASU CITY PD

Operator: M. MALOY JR #281
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	17:16:08
0.100 Cal Check	0.098	17:16:29
Air Blank	0.000	17:16:59

Preventative Maintenance
Performed -- Timer Reset